

Integrating the Healthcare Enterprise



5

IHE Devices (DEV) White Paper

10

Infusion Pump IV Flush

15

Revision 1.1 – Published

20

Date: May 9, 2025

Author: Devices Technical Committee / PCD Program / Infusion Pump WG

Email: dev@ihe.net

25

Please verify you have the most recent version of this document. See [here](#) for Published Versions and [here](#) for Public Comment versions.

Foreword

This is a white paper of the IHE Devices domain.

- 30 This white paper is published on May 9, 2025. Comments are invited and can be submitted at https://www.ihe.net/DEV_Public_Comments.

General information about IHE can be found at [IHE.net](https://www.ihe.net).

Information about the IHE Devices domain can be found at [IHE Domains](#).

- 35 Information about the organization of IHE Technical Frameworks and Supplements and the process used to create them can be found at [Profiles](#) and [IHE Processes](#).

The current version of the IHE Devices Technical Framework can be found at [Devices Technical Framework](#).

40

CONTENTS

1	Introduction	5
	1.1 Purpose of the Infusion Pump IV Flush White Paper.....	5
45	1.2 Intended Audience	5
2	Problem Description.....	5
3	Conventions	6
4	Definitions.....	7
5	Assumptions and Considerations	8
50	6 Use Cases	8
	6.1 Manual Programming Flush After Infusion Complete	8
	6.1.1 Description.....	8
	6.1.2 Precondition(s).....	8
	6.1.3 Use Case Flow	8
55	6.2 Manual Programming Flush After Syringe Empty	10
	6.2.1 Description.....	10
	6.2.2 Precondition(s).....	10
	6.2.3 Use Case Flow	11
	6.3 Manual Programming Flush After Micro Dose Volume Primed.....	14
60	6.3.1 Description.....	14
	6.3.2 Precondition(s).....	14
	6.3.3 Use Case Flow	14
	6.4 Manual Programming Flush with Primary After Secondary Completes	17
	6.4.1 Description.....	17
65	6.4.2 Precondition(s).....	17
	6.4.3 Main Flow.....	17
	6.5 Auto Programming Flush After Infusion Complete	20
	6.5.1 Description.....	20
	6.5.2 Precondition(s).....	20
70	6.5.3 Use Case Flow	20

	6.6 Auto Programming Flush After Syringe Empty	23
	6.6.1 Description.....	23
	6.6.2 Precondition(s).....	23
	6.6.3 Use Case Flow	23
75	6.7 Auto Programming Flush After Micro Volume Primed	27
	6.7.1 Description.....	27
	6.7.2 Precondition(s).....	27
	6.7.3 Use Case Flow	27
	Appendix A – Event Messages in HL7 format	31
80	A.1 Manual Flush After Infusion Complete	31
	A.1.1 Flush Start Event	31
	A.1.2 Flush Complete Event	33
	A.2 Manual Programming Flush After Syringe Empty	36
	A.2.1 Infusion Stop Event	36
85	A.2.2 Infusion Start Event (with two steps)	39
	A.2.3 Infusion Stop Event (with transitioning step).....	42
	A.2.4 Flush Start Event (Immediately following Stop Event)	45
	A.2.5 Flush Complete Event	48
	A.3 Auto Programming Flush After Infusion Complete.....	51
90	A.3.1 Flush Start Event	51
	A.3.2 Flush Complete Event	53
	A.4 Auto Programming Flush After Syringe Empty	55
	A.4.1 Infusion Start Event.....	55
	A.4.2 Infusion Complete Event.....	57
95	A.4.3 Flush Start Event	60
	A.4.4 Flush Complete Event	62
	Appendix B – Auto-Programming Request (APR) in HL7 format	64
	B.1 FLUSH APR PRN.....	64
100	B.2 FLUSH APR.....	64

1 Introduction

105 This document, the IHE Devices Infusion Pump IV Flush White Paper, describes the use of flushes as performed as a part of infusions. Health care providers use IV flushes to clear out IV lines that deliver medicine directly into the veins of a patient.

1.1 Purpose of the Infusion Pump IV Flush White Paper

This white paper is intended to identify a means for communicating flush events that are performed as a part of infusions. Flushes are administered before and after starting IV medication administrations. This has multiple effects such as:

- 110
- Ensures complete delivery of the ordered or prescribed medication.
 - Prevent adverse reactions between different medication administrations.
 - Prevent blockages and keep IV lines clean.

115 The primary goal of this paper is to define a solution that allows for the electronic transfer of infusion parameters and communicate the infusion pump events that are specifically related to flush use cases.

1.2 Intended Audience

The intended audience of the IHE Devices Infusion Pump IV Flush White Paper is:

- 120
- Clinicians involved in the use of infusion pumps in an integrated environment and others interested in integrating healthcare information systems and infusion pump workflows.
 - Experts involved in profile development from healthcare institutions, device and system manufacturers, and regulatory agencies.

2 Problem Description

125 To support current clinical workflows, this flush concept must meet several key user needs. It must allow the clinician to program the pump as the order is written, even when the line has been primed with medication. It must allow a flush syringe or primary infusate source to be easily programmed for the remaining line volume and with the same medication delivery rate. It must document the medication delivered and the flush phase as a single medication delivery (separate dose vs flush for HIS). Finally, this concept must support the interoperability workflow without a manual entry via automated programming requests from the HIS and allows the infusion pump to display what is infusing (drug or flush). These flush use case concepts must allow for communicating flush events, associating flush events with the related medication infusion, and allow for the auto-programming of those events to maintain seamless interoperability for all workflows between infusion pumps and hospital networks.

130

3 Conventions

135 For each event in the use cases (Section 5) all information is reported to a Health Information System (HIS) following IHE PCD-01 and PCD-10. The following information is reported to a HIS for each event:

- Infusion Event
- Pump Status
- 140 • Delivery Status
- Order ID
- Infusate Name
- Concentration (xx/mL):
 - Drug Amount
 - 145 ○ Dilute Volume (mL)
- Rate (mL/h)
- VTBI (mL)
- Volume Infused (mL)

150 For all auto programming requests (APRs) outlined in the use cases (Section 5) is reported to a HIS following IHE PCD-03. The following information is reported as a part of each APR:

- Clinician ID
- Patient ID
- Order Type
- Order ID
- 155 • Flush Solution or Medication
- Rate (mL/h)
- VTBI, if applicable

4 Definitions

Term	Definition	Acronym	Synonym
Auto Programming Request	Also known as a PCD-03 Communicate Infusion Order	APR	
Patient's Electronic Medical Administration Record	An electronic patient record system used to document information about medications administered to a patient.	eMAR	
Hospital Information System	A comprehensive, integrated information system that manages the hospital's operation and provides the eMAR along with partnering with the pump's interoperable ecosystem.	HIS	
Order/Med Order	Medication or fluid prescription to be delivered to a patient		
Pump Interoperability Gateway	Partner that interacts with IHE_HL7 compliant HIS systems to ensure both the outputs and the inputs of the system are compatible with IHE_HL7 systems.	PIG	
Patient Care Devices	A Work Program of the IHE Devices domain.	PCD	
PCD-01	IHE PCD transaction assigned to the Device Enterprise Communication(DEC) profile that is used to communicate patient care device data between systems.	DEC	
PCD-03	IHE PCD transaction assigned to the Point-of-Care Infusion Verification (PIV) profile that is used to communicate infusion order parameters between the device and the HIS	PIV	
PCD-10	IHE PCD transaction assigned to the Infusion Pump Event Communications (IPEC) profile that is used to communicate infusion event data between the infusion pump and the HIS	IPEC	
Pro Re Nata	Medication that should be administered as needed and whose timing is left to the patient, nurse or caregiver.	PRN	
Volume to Be Infused	The volume of fluid to be administered to a patient during an infusion or infusion step.	VTBI	

5 Assumptions and Considerations

The following assumptions and considerations are made for all the following flush use cases.

- Intermittent Dose Volume is equal to the diluent volume. Flush (PRN) Orders and Med Orders are not linked in the HIS. This is a standard configuration.
- The flush order and Med Order being linked is a rare configuration.
- The volume infused is the volume that the pump has pushed and is not the same as the volume that the patient has received.

6 Use Cases

6.1 Manual Programming Flush After Infusion Complete

6.1.1 Description

The disposable tubing is pre-primed with 0.5 mL of a saline solution. The syringe disposable container is loaded into the pump and the medication (Clindamycin) begins infusing. As the infusion completes, all the dose volume was infused into the tubing but there is still 0.5mL of medication left in the tubing waiting to go to the patient.

6.1.2 Precondition(s)

- The Dose Volume is less than or equal to the volume in the disposable.
- The Clindamycin infusion is back associated with the patient chart in the eMAR and is nearing completion.

6.1.3 Use Case Flow

1. The pump completes the VTBI and issues an Infusion Complete alarm.
2. The pump stops infusing and the PIG reports an infusion program Delivery Complete event to the HIS as follows:

<u>Reporting Data</u>	Active Primary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_COMP
<i>Pump Status:</i>	pump-status-not-infusing
Primary Channel Info	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason</i>	pump-stopped-standby
<i>Order ID:</i>	NULL
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	2.0000

RATE (mL/h):	1.3333
VTBI (mL):	2.0000
Vol. Infused (mL):	2.0000

3. The Clinical responds to the Infusion Complete alarm at the bedside.
4. The Clinical opens the EMR and following the BCMA process, scans the patient and the flush syringe disposable.
5. The Clinicians signs administration of flush in the eMAR.
6. The Clinician turns to the pump.
7. The Clinician unloads the medication syringe on the pump, connects and loads the new flush syringe containing about 1 mL of Normal Saline.
8. The Clinician adds a flush option and programs 0.9 mL for the VTBI.
9. The Clinician starts the flush step for Clindamycin on the pump.
10. The pump starts the Clindamycin flush and PIG reports to the HIS, a flush start event as follows (See Appendix A.1.1 for HL7 format).

<u>Reporting Data</u>	Active Flush Step
Infusion Event:	MDC_EVT_PUMP_DELIV_START
Pump Status	pump-status-infusing
Primary Channel Info	
Delivery Status:	pump-delivery-status-flushing
Med Order ID:	NULL
Flush Order ID:	NULL
Infusate Name:	Unknown
Drug Amt.:	----
Dil. Vol. (mL):	----
RATE (mL/h):	1.3333
VTBI (mL):	0.9000
Vol. Infused (mL):	0.0000
Flush Info	
VTBI (mL):	0.9000
Vol. Infused (mL):	0.0000

NOTE: Clinician does not manually program infusate name for flush. Hence, it is reported as “Unknown.”

11. When the VTBI of 0.9 mL is reached, the pump stops infusing and the PIG reports a flush complete event as follows (See Appendix A.1.2 for HL7 format):

200

Reporting Data		Active Flush Step
Infusion Event:	MDC_EVT_PUMP_DELIV_STOP	
Pump Status:	pump-status-not-infusing	
Primary Channel Info		
Delivery Status:	pump-delivery-status-not-delivering	
Not Delivering Reason	pump-stopped-flushing	
Med Order ID:	NULL	
Flush Order ID:	NULL	
Infusate Name:	Unknown	
Drug Amt.:	----	
Dil. Vol. (mL):	----	
RATE (mL/h):	1.3333	
VTBI (mL):	0.9000	
Vol. Infused (mL):	0.9000	
Flush Info		
VTBI (mL):	0.9000	
Vol. Infused (mL):	0.0000	

205

12. The Clinician reviews the I/O chart in the eMAR and signs values for 2 mL of Clindamycin and 0.9 mL of “Generic flush” over the 2.5 hours. Depending on the HIS implementation, the generic flush may be associated with the Clindamycin.

6.2 Manual Programming Flush After Syringe Empty

6.2.1 Description

210

The disposable tubing is primed with about 0.5 mL of medication from the 2 mL of the syringe disposable. This creates a situation in which the available volume in the medication syringe is less than what is needed to be infused. The medication disposable is then load into the pump and the remaining medication is administered. The infusion program results in medication container running to syringe empty, dose volume remaining to be infused and drug left in the tubing waiting to go to the patient.

6.2.2 Precondition(s)

215

- The Dose Volume is greater than the volume in the disposable.
- The Clindamycin infusion is associated with patient chart in the eMAR and nearing syringe empty.

6.2.3 Use Case Flow

1. Before completing the programmed VTBI, the pump detects the bottom of the disposable and issues a Syringe Empty alarm.
2. The pump stops and the PIG reports an infusion program stop event to the HIS as follows (See Appendix A.2.1 for HL7 format):

<u>Reporting Data</u>	Active Primary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_STOP
<i>Pump Status:</i>	pump-status-not-infusing
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason:</i>	pump-stopped-alarming
<i>Order ID:</i>	MED0101
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	2.0000
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	2.0000
<i>Vol. Infused (mL):</i>	1.5121

3. The Clinical responds to the Syringe Empty alarm at the bedside.
4. The Clinician opens EMR and following BCMA process, scans patient and the flush syringe disposable.
5. The Clinicians signs administration of flush in the eMAR.
6. The Clinician turns to the pump.
7. The Clinician unloads the empty medication syringe on the pump, connects and loads the new flush syringe containing about 1 mL of Normal Saline.
8. The Clinician adds a flush option and an additional flush volume of 0.9 mL to the VTBI.
9. The Clinician restarts the infusion program for Clindamycin on the pump.
10. The pump resumes the Clindamycin infusion and the PIG reports to the HIS, a start infusion event with two steps as follows (See Appendix A.2.2 for HL7 format):

<u>Reporting Data</u>	Active Primary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Source-Primary-Chan	
<i>Delivery Status:</i>	pump-delivery-status-delivering

Order ID:	MED0101
Infusate Name:	Clindamycin
Drug Amt.:	200.0000 mg
Dil. Vol. (mL):	2.0000
RATE (mL/h):	1.3333
VTBI (mL):	2.0000
Vol. Infused (mL):	1.5121
Pump-Flush-Info	
Infusate Name:	Unknown
VTBI (mL):	0.4121
Vol. Infused (mL):	0.0000

11. When the pump completes the original Clindamycin dose volume of 2 mL, the PIG reports to the HIS:

240

- a. The Clindamycin complete event is reported to the HIS as shown below (See Appendix A.2.3 for HL7 format). The complete event is reported as a Delivery Complete with status of transitioning.

<u>Reporting Data</u>	Active Primary Step
Infusion Event:	MDC_EVT_PUMP_DELIV_COMP
Pump Status:	pump-status-not-infusing
Infusate-Source-Primary-Chan	
Delivery Status:	pump-delivery-status-transitioning
Order ID:	MED0101
Infusate Name:	Clindamycin
Drug Amt.:	200.0000 mg
Dil. Vol. (mL):	2.0000 mL
RATE (mL/h):	1.3333
VTBI (mL):	2.0000
Vol. Infused (mL):	2.0000
Pump-Flush-Info	
Infusate Name:	Unknown
VTBI (mL):	0.4121
Vol. Infused (mL):	0.0000

245

- b. The complete event is immediately followed by the flush start event as follows (See Appendix A.2.4 for HL7 format):

<u>Reporting Data</u>	Active Flush Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Source-Primary-Chan	
<i>Delivery Status:</i>	pump-delivery-status-flushing
<i>Med Order ID:</i>	MED0101
<i>Flush Order ID:</i>	NULL
<i>Infusate Name:</i>	Unknown
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	0.4121
<i>Vol. Infused (mL):</i>	0.0000
Pump-Flush-Info	
<i>VTBI (mL):</i>	0.4121
<i>Vol. Infused (mL):</i>	0.0000

250

12. When the additional VTBI of 0.9 mL is reached, the pump stops, and the PIG reports a flush complete event as follows (See Appendix A.2.5 for HL7 format):

<u>Reporting Data</u>	Active Flush Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_STOP
<i>Pump Status:</i>	pump-status-not-infusing
Infusate-Source-Primary-Chan	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason:</i>	pump-stopped-flushing
<i>Med Order ID:</i>	MED0101
<i>Flush Order ID:</i>	NULL
<i>Infusate Name:</i>	unknown
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	0.4121
<i>Vol. Infused (mL):</i>	0.4121
Pump-Flush-Info	
<i>VTBI (mL):</i>	0.4121
<i>Vol. Infused (mL):</i>	0.0000

- 255 13. The Clinician reviews I/O chart in the eMAR and signs values for 2 mL of Clindamycin and 0.4121 mL of “Generic flush” over the last 2 hours. The generic flush is associated with Clindamycin.

6.3 Manual Programming Flush After Micro Dose Volume Primed

6.3.1 Description

260 The disposable tubing is primed with small volume medication (dose vol. < tubing volume) and requires immediate start of flush disposable to deliver medication. As the infusion completes, all dose volume is infused to the patient and no medication left in tubing.

6.3.2 Precondition(s)

- The Dose Volume (2mL) is less than the tubing volume (3mL).
- The Dose Volume is manually pushed into the disposable tubing between the pump and the patient.

265 6.3.3 Use Case Flow

1. The Clinician opens the EMR and following the BCMA process, scans the patient and the medication disposable.
2. The Clinicians signs administration of medication in the eMAR.
3. The Clinician turns to the pump.
- 270 4. After pushing all the medication into the tubing below the pump, the Clinician loads a flush syringe (3mL or greater) into the pump.
5. The Clinician selects Clindamycin from the pump library and proceeds to program the medication parameters.
6. The Clinician reviews the program and starts infusion.
- 275 7. The pump starts infusing and the PIG reports a start event as follows:

<u>Reporting Data</u>	Active Primary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Source-Primary-Chan	
<i>Delivery Status:</i>	pump-delivery-status-delivering
<i>Order ID:</i>	NULL
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	2.0000
<i>RATE (mL/h):</i>	1.3333

VTBI (mL):	2.0000
Vol. Infused (mL):	0.0000

8. The Clinician back associates the infusion with the order in the eMAR.
9. When the programmed VTBI of 2mL is reached, the pump stops infusing and the PIG reports an infusion program Delivery Complete event as follows:

<u>Reporting Data</u>	Active Primary Step
Infusion Event:	MDC_EVT_PUMP_DELIV_COMP
Pump Status:	pump-status-not-infusing
Infusate-Source-Primary-Chan	
Delivery Status:	pump-delivery-status-not-delivering
Not Delivering Reason	pump-stopped-standby
Order ID:	NULL
Infusate Name:	Clindamycin
Drug Amt.:	200.0000 mg
Dil. Vol. (mL):	2.0000
RATE (mL/h):	1.3333
VTBI (mL):	2.0000
Vol. Infused (mL):	2.0000

10. The Clinician responds to the Infusion Complete alarm at the bedside.
11. Clinician reviews I/O chart in EMAR and signs value for 2 mL of Clindamycin in patient chart.
12. The Clinician following BCMA process, scans patient and the flush syringe disposable.
13. The Clinicians signs administration of flush in the eMAR.
14. The Clinician turns to the pump.
15. The Clinician proceeds with the flush programming and enters 1 mL as the VTBI and selects START.
16. The pump starts infusing and the IE reports an infusion start event as follows:

<u>Reporting Data</u>	Active Flush Step
Infusion Event:	MDC_EVT_PUMP_DELIV_START
Pump Status:	pump-status-infusing
Infusate-Source-Primary Chan	
Delivery Status:	pump-delivery-status-flushing

<i>Med Order ID:</i>	NULL
<i>Flush Order ID:</i>	NULL
<i>Infusate Name:</i>	unknown
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	1.0000
<i>Vol. Infused (mL):</i>	0.0000
Pump-Flush-Info	
<i>VTBI (mL):</i>	1.0000
<i>Vol. Infused (mL):</i>	0.0000

- 295 17. When the VTBI of 1 mL is reached, the pump stops infusing and the PIG reports a flush complete event as follows:

<u>Reporting Data</u>	<u>Active Flush Step</u>
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_STOP
<i>Pump Status:</i>	pump-status-not-infusing
Infusate-Source-Primary-Chan	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason</i>	pump-stopped-flushing
<i>Med Order ID:</i>	NULL
<i>Flush Order ID:</i>	NULL
<i>Infusate Name:</i>	unknown
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	1.0000
<i>Vol. Infused (mL):</i>	1.0000
Pump-Flush-Info	
<i>VTBI (mL):</i>	1.0000
<i>Vol. Infused (mL):</i>	0.0000

- 300 18. Clinician reviews I/O chart in eMAR and signs values for 2 mL of Clindamycin and 1 mL of “Generic flush”. Depending on HIS implementation, the generic flush may be associated with the Clindamycin.

6.4 Manual Programming Flush with Primary After Secondary Completes

6.4.1 Description

305 Disposable tubing is primed with fluid from primary infusate source. The medication is then delivered from a secondary/piggyback source. The infusion program results in the medication container running to dose volume complete and medication volume left in the tubing waiting to go to the patient.

6.4.2 Precondition(s)

- The Dose Volume is equal to the volume in the disposable.
- 310 • The Clindamycin infusion is associated with patient chart in eMAR.

6.4.3 Main Flow

1. Program and start IV fluid.

<u>Reporting Data</u>	<u>Active Primary Step</u>
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-delivering
<i>Order ID:</i>	MED0101
<i>Infusate Name:</i>	IV Fluid
<i>Drug Amt.:</i>	-
<i>Dil. Vol. (mL):</i>	-
<i>RATE (mL/h):</i>	1.0000
<i>VTBI (mL):</i>	100.0000
<i>Vol. Infused (mL):</i>	0

- 315 2. Program prescribed medication on secondary line (IVPB).
3. Program flush volume of 5 mL to deliver following IVPB.
4. Start IVPB.
5. The pump stops the primary infusion line and the PIG reports to the HIS, an infusion program stop events as follows:

320

<u>Reporting Data</u>	<u>Active Primary Step</u>
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_STOP
<i>Pump Status:</i>	pump-status-not-infusing

Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-switching-source
<i>Order ID:</i>	MED0101
<i>Infusate Name:</i>	IV Fluid
<i>Drug Amt.:</i>	-
<i>Dil. Vol. (mL):</i>	-
<i>RATE (mL/h):</i>	1.0000
<i>VTBI (mL):</i>	100.0000
<i>Vol. Infused (mL):</i>	1.5000

6. The pump starts the Clindamycin infusion on the secondary line and the PIG reports to the HIS, an infusion program start event as follows:

Reporting Data	Active Secondary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Pump-Secondary-Info	
<i>Delivery Status:</i>	pump-delivery-status-delivering
<i>Order ID:</i>	MED0102
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	10.0000
<i>RATE (mL/h):</i>	15.0000
<i>VTBI (mL):</i>	10.0000
<i>Vol. Infused (mL):</i>	0

- 325 7. When the pump completes the Clindamycin dose volume of 10 mL on the secondary line, the PIG reports to the HIS:
- a. The Clindamycin Delivery Complete event as follows:

Reporting Data	Active Secondary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_COMP
<i>Pump Status:</i>	pump-status-not-infusing
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
Pump-Secondary-Info	
<i>Order ID:</i>	MED0101
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	10.0000
<i>RATE (mL/h):</i>	15.0000
<i>VTBI (mL):</i>	10.0000

Vol. Infused (mL):	10.0000
---------------------------	---------

330

- b. Immediately followed by transition of the pump back to the primary line, the PIG reports to the HIS, a Start Event for Clindamycin flush follows. The volume infused is accumulated from the primary bag:

Reporting Data	Active Flush Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-flushing
<i>Med Order ID:</i>	MED0101
<i>Flush Order ID:</i>	NULL
<i>Infusate Name:</i>	IV Fluid
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	15.0000
<i>VTBI (mL):</i>	5.0000
<i>Vol. Infused (mL):</i>	1.5000
Pump-Flush-Info	
<i>VTBI (mL):</i>	5.0000
<i>Vol. Infused (mL):</i>	0.0000

335

8. When the programmed VTBI of 5 mL is reached, the pump stops the flush and the PIG reports a flush complete event as follows:

Reporting Data	Active Flush Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_COMP
<i>Pump Status:</i>	pump-status-not-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason</i>	pump-stopped-flushing
<i>Med Order ID:</i>	MED0101
<i>Flush Order ID:</i>	NULL
<i>Infusate Name:</i>	IV Fluid
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	15.0000
<i>VTBI (mL):</i>	5.0000
<i>Vol. Infused (mL):</i>	6.5000
Pump-Flush-Info	

VTBI (mL):	5.0000
Vol. Infused (mL):	5.0000

9. The pump automatically resumes the original primary rate on the primary line:

340

Reporting Data	Active Primary Step
Infusion Event:	MDC_EVT_PUMP_DELIV_START
Pump Status:	pump-status-infusing
Infusate-Primary-Source-Chan	
Delivery Status:	pump-delivery-status-delivering
Order ID:	MED0101
Infusate Name:	IV Fluid
Drug Amt.:	-
Dil. Vol. (mL):	-
RATE (mL/h):	1.0000
VTBI (mL):	100.0000
Vol. Infused (mL):	6.5000

6.5 Auto Programming Flush After Infusion Complete

6.5.1 Description

345

Disposable tubing is pre-primed with 0.5 mL of a solution. The syringe disposable is load into the pump and the medication begins infusing. The medication step results in medication container running to infusion complete, all the dose volume was infused by pump but there is medication volume left in the tubing waiting to go to the patient.

6.5.2 Precondition(s)

350

- A PRN flush order exists in the HIS
- The Dose Volume is less than the volume in the disposable.
- The Clindamycin infusion was associated with patient chart in eMAR and nearing completion.

6.5.3 Use Case Flow

355

1. The pump completes the VTBI and issues an Infusion Completion alarm.
2. The pump stops infusing and the PIG reports to the HIS, an infusion program Delivery Complete event as follows:

<u>Reporting Data</u>	Active Primary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_COMP
<i>Pump Status:</i>	pump-status-not-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason</i>	pump-stopped-standby
<i>Order ID:</i>	NULL
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	2.0000
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	2.0000
<i>Vol. Infused (mL):</i>	2.0000

3. The Clinician responds to the Infusion Complete alarm at the bedside.
- 360 4. The Clinician opens EMR and following the BCMA process, scan the patient, flush syringe disposable, and the pump.
5. The Clinician sends a FLUSH APR PRN order for Normal Saline to the pump as follows (See Appendix B.1 for APR HL7 format):

Flush APR Details	
<i>Clinician ID:</i>	RatchedRN
<i>Patient ID:</i>	W.Yueh
<i>Order Type:</i>	PrnFlush
<i>Order ID:</i>	FLUSHIE2000
<i>Flush Solution:</i>	Normal Saline
<i>VTBI.:</i>	0.5000

- 365 6. The PIG validates and accepts the Flush Request.
7. The Clinician turns to the pump.
8. The Clinician unloads the medication syringe and connects and loads the flush syringe containing about 2 mL of Normal Saline.
- 370 9. The Clinician reviews auto-populated parameters and starts the flush step for Clindamycin on the pump.
10. The pump starts the Clindamycin flush and PIG reports to the HIS, a flush start event as follows (See Appendix A.3.1 for HL7 format):

<u>Reporting Data</u>	Active Flush Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-flushing
<i>Med Order ID:</i>	NULL
<i>Flush Order ID:</i>	FLUSHIE2000
<i>Infusate Name:</i>	Normal Saline
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	0.5000
<i>Vol. Infused (mL):</i>	0.0000
Pump-Flush-Info	
<i>VTBI (mL):</i>	0.5000
<i>Vol. Infused (mL):</i>	0.0000

375

11. The Clinician returns to the eMAR to sign the administration of the flush.
12. When the VTBI of 0.5 mL from the APR is reached, the pump stops infusing and the PIG reports a flush complete event as follows (See Appendix A.3.2 for HL7 format):

<u>Reporting Data</u>	Active Flush Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_STOP
<i>Pump Status:</i>	pump-status-not-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason</i>	pump-stopped-flushing
<i>Med Order ID:</i>	NULL
<i>Flush Order ID:</i>	FLUSHIE2000
<i>Infusate Name:</i>	Normal Saline
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	0.5000
<i>Vol. Infused (mL):</i>	0.5000
Pump-Flush-Info	
<i>VTBI (mL):</i>	0.5000

Vol. Infused (mL):	0.5000
---------------------------	--------

380

13. Clinician reviews I/O chart in eMAR and signs values for 2 mL of Clindamycin and 0.5 mL of Normal Saline infused over the last 2 hours. Depending on HIS implementation, the saline flush may be associated with the Clindamycin.

6.6 Auto Programming Flush After Syringe Empty

6.6.1 Description

Disposable tubing is primed with about 0.5 mL of medication from the 2 mL of the syringe disposable. This creates a situation in which the available volume in the medication syringe is less than what is needed to be infused. The medication disposable is then loaded into the pump and the remaining medication is administered. The infusion program results in medication container running to syringe empty, dose volume remaining to be infused and drug left in the tubing waiting to go to the patient.

390

6.6.2 Precondition(s)

- A PRN flush order exists in the HIS
- The Dose Volume is greater than the volume in the disposable.
- The Clindamycin infusion is associated with patient chart in eMAR and nearing syringe empty.

395

6.6.3 Use Case Flow

1. Before completing the programmed VTBI, the pump detects the bottom of the disposable and issues a Syringe Empty alarm.
2. The pump stops and the PIG reports to the HIS, an infusion program stop event as follows:

400

<u>Reporting Data</u>	Active Primary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_STOP
<i>Pump Status:</i>	pump-status-not-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason:</i>	pump-stopped-alarming
<i>Order ID:</i>	MED0101
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	2.0000

RATE (mL/h):	1.3333
VTBI (mL):	2.0000
Vol. Infused (mL):	1.5121

3. The Clinician responds to the Syringe Empty alarm at the bedside.
4. The Clinician opens EMR and following BCMA process, scans patient, flush syringe disposable, and the pump.
5. The Clinician sends a FLUSH APR PRN order for Normal Saline to the pump as follows (See Appendix B.1 for HL7 format):

Flush APR Details	
<i>Clinician ID:</i>	RatchedRN
<i>Patient ID:</i>	W.Yueh
<i>Order Type:</i>	PrnFlush
<i>Flush Order ID:</i>	FLUSHIE2000
<i>Flush Solution:</i>	Normal Saline
<i>VTBI:</i>	1.0000

6. The PIG validates and accepts the Flush Request.
7. The Clinician turns to the pump.
8. The Clinician unloads the empty medication syringe on the pump, connects and loads the scanned flush syringe containing about 1 mL of Normal Saline.
9. The Clinician reviews auto-populated parameters and restarts the Clindamycin infusion.
10. The pump resumes the Clindamycin infusion and PIG reports to the HIS, a start infusion event as follows (See Appendix A.4.1 for HL7 format):

Reporting Data	Active Primary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-delivering
<i>Order ID:</i>	MED0101
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	2.0000

RATE (mL/h):	1.3333
VTBI (mL):	2.0000
Vol. Infused (mL):	1.5121
Pump-Flush-Info	
Pump Status:	pump-status-not-infusing
Delivery Status:	pump-stopped-standby
Infusate Name:	Normal Saline
VTBI (mL):	0.5121
Vol. Infused (mL):	0.0000

- 420 11. The Clinician returns to the eMAR to sign the administration of the flush.
12. When the Clindamycin volume infused reaches 2 mL which is equal to the dose volume, the PIG reports to the HIS:
- a. The Clindamycin Delivery Complete event as follows (See Appendix A.4.2 for HL7 format):

425

<u>Reporting Data</u>	Active Primary Step
Infusion Event:	MDC_EVT_PUMP_DELIV_COMP
Pump Status:	pump-status-not-infusing
Infusate-Primary-Source-Chan	
Delivery Status:	pump-delivery-status-transitioning
Order ID:	MED0101
Infusate Name:	Clindamycin
Drug Amt:	200.0000 mg
Dil. Vol. (mL):	2.0000 mL
RATE (mL/h):	1.3333
VTBI (mL):	2.0000
Vol. Infused (mL):	2.0000
Pump-Flush-Info	
Infusate Name:	Normal Saline
VTBI (mL):	0.5121
Vol. Infused (mL):	0.0000

- b. Immediately followed by a start event for the Clindamycin flush as follows (See Appendix A.4.3 for HL7 format):

<u>Reporting Data</u>	Active Flush Step
------------------------------	--------------------------

<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-flushing
<i>Med Order ID:</i>	MED0101
<i>Flush Order ID:</i>	FLUSHIE2000
<i>Infusate Name:</i>	Normal Saline
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	0.5121
<i>Vol. Infused (mL):</i>	0.0000
Pump-Flush-Info	
<i>VTBI (mL):</i>	0.5121
<i>Vol. Infused (mL):</i>	0.0000

430

13. When the VTBI of 1 mL from the APR is reached, the pump stops infusing and the PIG reports a flush complete event (See Appendix A.4.4 for HL7 format):

<u>Reporting Data</u>	Active Flush Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_STOP
<i>Pump Status:</i>	pump-status-not-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason</i>	pump-stopped-flushing
<i>Med Order ID:</i>	MED0101
<i>Flush Order ID:</i>	FLUSHIE2000
<i>Infusate Name:</i>	Normal Saline
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	0.5121
<i>Vol. Infused (mL):</i>	0.5121
Pump-Flush-Info	
<i>VTBI (mL):</i>	0.5121
<i>Vol. Infused (mL):</i>	0.5121

- 435 14. The Clinician reviews I/O chart in eMAR and signs values for 2 mL of Clindamycin and 0.5121 mL of Normal Saline infused. The saline flush is associated with the Clindamycin.

6.7 Auto Programming Flush After Micro Volume Primed

6.7.1 Description

- 440 Disposable tubing is primed with small volume medication (dose vol. < tubing volume) and requires immediate start of flush disposable. The flush results in disposable running to infusion complete, all dose volume infused, and no medication left in tubing.

6.7.2 Precondition(s)

- The Dose Volume is less than the tubing volume.
- 445 • Dose volume is manually pushed into tubing between pump and patient.

6.7.3 Use Case Flow

1. The Clinician opens EMR and following BCMA process, scans patient, medication syringe disposable, and the pump.
2. The Clinician sends a medication APR for Clindamycin to the Pump
- 450 3. The Clinician turns to the pump.
4. After pushing all the medication into the tubing below the pump, the Clinician loads Flush disposable into the pump.
5. The Clinician reviews auto-populated parameters and starts the Clindamycin infusion.
6. The pump starts infusing and the PIG reports a start event on the pump as follows:

455

<u>Reporting Data</u>	Active Primary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-delivering
<i>Order ID:</i>	NULL
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	2.0000
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	2.0000
<i>Vol. Infused (mL):</i>	0.0000

7. The Clinician returns to the eMAR to sign the administration of the Clindamycin medication.
8. When the programmed VTBI of 2 mL is reached, the pump stops infusing and the PIG reports an infusion program Delivery Complete event as follows:

<u>Reporting Data</u>	Active Primary Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_COMP
<i>Pump Status:</i>	pump-status-not-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason</i>	pump-stopped-standby
<i>Order ID:</i>	NULL
<i>Infusate Name:</i>	Clindamycin
<i>Drug Amt.:</i>	200.0000 mg
<i>Dil. Vol. (mL):</i>	2.0000
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	2.0000
<i>Vol. Infused (mL):</i>	2.0000

9. The Clinician responds to the Infusion Complete alarm at the bedside.
10. The Clinician reviews I/O chart in eMAR and sees a row for 2 mL of Clindamycin in patient chart.
11. The Clinician opens EMR and following BCMA process, scans patient, flush syringe disposable, and the pump.
12. Clinician sends a FLUSH APR order for Normal Saline to the Pump as follows (See Appendix B.2 for HL7 format):

Flush APR Details	
<i>Clinician ID:</i>	RatchedRN
<i>Patient ID:</i>	W.Yueh
<i>Order Type:</i>	PrnFlush
<i>Order ID:</i>	FLUSHIE2000
<i>Flush Solution:</i>	Normal Saline
<i>VTBI.:</i>	0.5000

13. The PIG validates and accepts the Flush Request.
14. The Clinician turns to the pump.

15. The Clinician reviews auto-populated parameters and starts the flush step for Clindamycin on the pump.

475

16. The pump starts the Clindamycin flush and PIG reports to the HIS, a flush start event as follows:

<u>Reporting Data</u>	Active Flush Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_START
<i>Pump Status:</i>	pump-status-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-flushing
<i>Med Order ID:</i>	NULL
<i>Flush Order ID:</i>	FLUSHIE2000
<i>Infusate Name:</i>	Normal Saline
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333
<i>VTBI (mL):</i>	0.5000
<i>Vol. Infused (mL):</i>	0.0000
Pump-Flush-Info	
<i>VTBI (mL):</i>	0.5000
<i>Vol. Infused (mL):</i>	0.0000

17. The Clinician returns to the eMAR to sign the administration of the flush.

480

18. When the VTBI of 0.5 mL is reached, the pump stops infusing and the PIG reports a flush complete event as follows:

<u>Reporting Data</u>	Active Flush Step
<i>Infusion Event:</i>	MDC_EVT_PUMP_DELIV_STOP
<i>Pump Status:</i>	pump-status-not-infusing
Infusate-Primary-Source-Chan	
<i>Delivery Status:</i>	pump-delivery-status-not-delivering
<i>Not Delivering Reason</i>	pump-stopped-flushing
<i>Med Order ID:</i>	NULL
<i>Flush Order ID:</i>	FLUSHIE2000
<i>Infusate Name:</i>	Normal Saline
<i>Drug Amt.:</i>	----
<i>Dil. Vol. (mL):</i>	----
<i>RATE (mL/h):</i>	1.3333

<i>VTBI (mL):</i>	0.5000
<i>Vol. Infused (mL):</i>	0.5000
Pump-Flush-Info	
<i>VTBI (mL):</i>	0.5000
<i>Vol. Infused (mL):</i>	0.5000

- 485 19. The Clinician reviews I/O chart in eMAR and signs values for 2 mL of Clindamycin and 0.5 mL of Normal Saline infused. The saline flush is associated with the Clindamycin.

Appendices

Appendix A – Event Messages in HL7 format

The key portions of the event messages are highlighted in yellow to demonstrate the differences between events as seen in the use cases.

A.1 Manual Flush After Infusion Complete

490 A.1.1 Flush Start Event

```

MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
PID||2392^^^HL7^PI||^U|||||
495 OBR|1|Unknown^HL7^00D075000000154^EUI-64|744^PUMPVENDOR^744^EUI-
64|999999^Medication_Unknown||20220517084831-0700|||||
OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0||||X||20220517084831-
0700||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0||||X||20220517084831-
500 0700||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197288^MDC_EVT_PUMP_DELIV_START^MDC||||R||202205170
84831-0700||||20220517084831-0700
OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2||||X||20220517084831-0700||||20220517084831-0700
OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0||||X||20220517084831-0700||||20220517084831-0700
505 OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-infusing^MDC||||R||20220517084831-
0700||||20220517084831-0700
OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-flush||||R||20220517084831-
0700||||20220517084831-0700
OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-flush||||R||20220517084831-
510 0700||||20220517084831-0700

```

OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.4|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||||R||20220517084831-0700||||20220517084831-0700

OBX|10|70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.8.0||||X||20220512135317-0700||||20220512135317-0700

515 OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.8.1|A||||R||20220517084831-0700||||20220517084831-0700

OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.8.1|^pump-delivery-status-flushing^MDC||||R||20220512135317-0700||||20220512135317-0700

520 OBX|13|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.8.2|^pump-program-delivery-mode-continuous^MDC||||R||20220512135317-0700||||20220512135317-0700

OBX|14|NM|157784^MDC_FLOW_FLUID_PUMP^MDC|1.1.8.3|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||||R||20220517091144-0700||||20220517091144-0700

OBX|15|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.4|Unknown||||R||20220512135317-0700||||20220512135317-0700

OBX|16|ST|184515^MDC_DRUG_ID^MDC|1.1.8.5|Unknown||||R||20220512135317-0700||||20220512135317-0700

525 OBX|17|NM|157993^MDC_VOL_FLUID_DELIV_TOTAL^MDC|1.1.8.6|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|18|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.7|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

530 OBX|19|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.8|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|20|70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700

OBX|21|ST|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700

OBX|22|ST|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-0700

535 OBX|23|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|24|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|26|0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700

540 OBX|30|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.10.4|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|31|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.10.5|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||2022
0512135317-0700||||20220512135317-0700
OBX|32|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.10.6|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCU
M||||R||20220512135317-0700||||20220512135317-0700

545

A.1.2 Flush Complete Event

MSH|^~\&| DEVICE_PUMPVENDOR ^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
PID||2392^^^HL7^PI||^U|||||||
OBR|1|Unknown^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|999999^Medication_Unknown||20220517084831-
0700|||||
OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0||||X||20220517084831-
0700||||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0||||X||20220517084831-
0700||||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197290^MDC_EVT_PUMP_DELIV_STOP^MDC||||R||2022051708
4831-0700||||20220517084831-0700
OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2||||X||20220517084831-0700||||20220517084831-0700
OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0||||X||20220517084831-0700||||20220517084831-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-not-infusing^MDC||||R||20220517084831-
0700||||20220517084831-0700
OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-flush||||R||20220517084831-
0700||||20220517084831-0700
OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-flush||||R||20220517084831-
0700||||20220517084831-0700
OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|0|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||
||R||20220517084831-0700||||20220517084831-0700

550

555

560

565

570 OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC^MDC|1.1.8.0|||||X||20220512135317-0700||||20220512135317-0700
OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.8.1|A|||||R||20220517084831-0700||||20220517084831-0700
OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.8.1|^pump-delivery-status-not-delivering^MDC|||||R||20220512135317-0700||||20220512135317-0700
575 OBX|13|CWE|158006^MDC_PUMP_NOT_DELIVERING_REASON^MDC|1.1.8.2|^pump-stopped-flush^MDC|||||R||20220517092221-0700||||20220517092221-0700
OBX|14|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.8.3|^pump-program-delivery-mode-continuous^MDC|||||R||20220512135317-0700||||20220512135317-0700
OBX|15|NM|157784^MDC_FLOW_FLUID_PUMP^MDC|1.1.8.9|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM|||||R||20220517091144-0700||||20220517091144-0700
580 OBX|16|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Unknown|||||R||20220512135317-0700||||20220512135317-0700
OBX|17|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|Unknown|||||R||20220512135317-0700||||20220512135317-0700
OBX|18|NM|157993^MDC_VOL_FLUID_DELIV_TOTAL^MDC|1.1.8.31|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700
585 OBX|19|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700
OBX|20|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700
OBX|21|ST|70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0|||||X||20220512135317-0700||||20220512135317-0700
590 OBX|22|ST|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10|||||R||20220512135317-0700||||20220512135317-0700
OBX|23|NM|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD|||||R||20220512135317-0700||||20220512135317-0700
OBX|24|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700
595 OBX|25|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700
OBX|26||0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0|||||X||20220512135317-0700||||20220512135317-0700

600 OBX|27|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.10.4|0.9000|263762^MDC_DIM_MILLI_L^MDC
^^mL^UCUM||||R|||20220512135317-0700||||20220512135317-0700
OBX|28|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.10.5|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R|||2022
0512135317-0700||||20220512135317-0700
OBX|29|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.10.6|0.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCU
M||||R|||20220512135317-0700||||20220512135317-0700

605

A.2 Manual Programming Flush After Syringe Empty

A.2.1 Infusion Stop Event

```

610 MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517092221-0700||ORU^R42^ORU_R01|8db9924f-5e2a-4e61-9f59-d53990da2d1c|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
PID||1658^^^HL7^PI||^U|||||
OBR|1|MED0101^HL7^00D0750000000165^EUI-64|1361^Neo^1361^EUI-64|Clindamycin^Clindamycin||20220517091144-
0700|||||
615 OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0||||X||20220517092221-
0700||||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517092221-0700
OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0||||X||20220517092221-
0700||||M8110S12345^SyrPump^00D0750036E58879^EUI-64|20220517092221-0700
OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197290^MDC_EVT_PUMP_DELIV_STOP^MDC||||R||2022051709
2221-0700||||20220517092221-0700
620 OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2||||X||20220517092221-0700||||20220517092221-0700
OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0||||X||20220517092221-0700||||20220517092221-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-not-infusing^MDC||||R||20220517092221-
0700||||20220517092221-0700
OBX|7|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-primary||||R||20220517092221-
625 0700||||20220517092221-0700
OBX|8|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|0.0000|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UC
UM||||R||20220517092221-0700||||20220517092221-0700
OBX|9|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A||||R||20220517092221-
0700||||20220517092221-0700
630 OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.2.0||||X||20220517092221-
0700||||20220517092221-0700
OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.2.1|^pump-delivery-status-not-
deliverng^MDC||||R||20220517091144-0700||||20220517091144-0700

```

635 OBX|13|CWE|158006^MDC_PUMP_NOT_DELIVERING_REASON^MDC|1.1.8.2|^pump-stopped-
alarming^MDC||||R||20220517092221-0700||||20220517092221-0700
OBX|14|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.2.3|^pump-program-delivery-mode-
continuous^MDC||||R||20220517091144-0700||||20220517091144-0700
OBX|15|NM|157784^MDC_FLOW_FLUID_PUMP^MDC|1.1.2.9|0|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM|
||||R||20220517091144-0700||||20220517091144-0700
640 OBX|16|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.2.10|Clindamycin||||R||20220517091144-0700||||20220517091144-
0700
OBX|17|SN|157760^MDC_CONC_DRUG^MDC|1.1.2.11|200.0000/2.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/m
L^UCUM||||R||20220517091144-0700||||20220517091144-0700
OBX|18|ST|184515^MDC_DRUG_ID^MDC|1.1.2.15|Clindamycin||||R||20220517091144-0700||||20220517091144-0700
645 OBX|19|ST|184516^MDC_PUMP_DRUG_LIBRARY_CARE_AREA^MDC|1.1.2.16|Some Drug Library||||R||20220517091144-
0700||||20220517091144-0700
OBX|20|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.2.31|1.5121|263762^MDC_DIM_MILLI_L^MDC
^^mL^UCUM||||R||20220517091144-0700||||20220517091144-0700
OBX|21|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.2.32|0.4879|263762^MDC_DIM_MILLI_L^MDC^^mL^UCU
M||||R||20220517091144-0700||||20220517091144-0700
650 OBX|22|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.2.33|2.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||2022
0517091144-0700||||20220517091144-0700
OBX|23|NM|157924^MDC_RATE_DOSE^MDC|1.1.2.41|200.0000|263890^MDC_DIM_MILLI_G^MDC^^mg^UCUM||||R||20220
517091144-0700||||20220517091144-0700
655 OBX|24|NM|157760^MDC_CONC_DRUG^MDC|1.1.2.111|100.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/mL^U
CUM||||R||20220517091144-0700||||20220517091144-0700
OBX|25||70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700
OBX|26|NM|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700
OBX|27|NM|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-
660 0700
OBX|28|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||202205
12135317-0700||||20220512135317-0700

OBX|29|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|0.9000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||
R|||20220512135317-0700||||20220512135317-0700

665

A.2.2 Infusion Start Event (with two steps)

MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVEDNOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
670 PID||2392^^^HL7^PI||^U|||||||
OBR|1|MED0101^HL7^00D075000000154^EUI-64|744^SyrPump^744^EUI-64|Clindamycin^Clindamycin||20220517084831-
0700|||||
OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0||||X||20220517084831-
0700|||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
675 OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0||||X||20220517084831-
0700|||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197288^MDC_EVT_PUMP_DELIV_START^MDC||||R||202205170
84831-0700||||20220517084831-0700
OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2||||X||20220517084831-0700||||20220517084831-0700
680 OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0||||X||20220517084831-0700||||20220517084831-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-infusing^MDC||||R||20220517084831-
0700||||20220517084831-0700
OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-primary~^pump-source-info-
flush||||R||20220517084831-0700||||20220517084831-0700
685 OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-primary||||R||20220517084831-
0700||||20220517084831-0700
OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UC
UM||||R||20220517084831-0700||||20220517084831-0700
OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.2.0||||X||20220517084831-
690 0700||||20220517084831-0700
OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A||||R||20220517084831-
0700||||20220517084831-0700
OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.2.1|^pump-delivery-status-
delivering^MDC||||R||20220517084831-0700||||20220517084831-0700

695 OBX|13|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.2.3|^pump-program-delivery-mode-continuous^MDC||||R||20220517084831-0700||||20220517084831-0700
OBX|14|NM|157784^MDC_FLOW_FLUID_PUMP^MDC|1.1.2.9|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||||R||20220517084831-0700||||20220517084831-0700
700 OBX|15|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.2.10|Clindamycin||||R||20220517084831-0700||||20220517084831-0700
OBX|16|SN|157760^MDC_CONC_DRUG^MDC|1.1.2.11|200.0000/2.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
OBX|17|ST|184515^MDC_DRUG_ID^MDC|1.1.2.15|Clindamycin||||R||20220517084831-0700||||20220517084831-0700
705 OBX|18|ST|184516^MDC_PUMP_DRUG_LIBRARY_CARE_AREA^MDC|1.1.2.16|Some Drug Library||||R||20220517084831-0700||||20220517084831-0700
OBX|19|ST|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.2.31|1.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
OBX|20|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.2.32|0.4879|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
710 OBX|21|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.2.33|2.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
OBX|22|NM|157924^MDC_RATE_DOSE^MDC|1.1.2.41|200.0000|263890^MDC_DIM_MILLI_G^MDC^^mg^UCUM||||R||20220517084831-0700||||20220517084831-0700
OBX|23|NM|157760^MDC_CONC_DRUG^MDC|1.1.2.111|100.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
715 OBX|24||70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700
OBX|25|NM|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.33|BD-10||||R||20220512135317-0700||||20220512135317-0700
OBX|26|NM|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.33|BD||||R||20220512135317-0700||||20220512135317-0700
720 OBX|27|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.33|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|28|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.33|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|29||0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700

725 OBX|33|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline||||R||20220512135317-0700||||20220512135317-0700
 OBX|34|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|Unknown||||R||20220512135317-0700||||20220512135317-0700
 OBX|35|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
 730 OBX|36|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
 OBX|37|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

A.2.3 Infusion Stop Event (with transitioning step)

MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
PID||2392^^^HL7^PI||^U|||||||
OBR|1|MED0101^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|Clindamycin^Clindamycin||20220517084831-
0700|||||
OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0||||X||20220517084831-
0700|||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0||||X||20220517084831-
0700|||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197290^MDC_EVT_PUMP_DELIV_COMP^MDC||||R||2022051708
4831-0700||||20220517084831-0700
OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2||||X||20220517084831-0700||||20220517084831-0700
OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0||||X||20220517084831-0700||||20220517084831-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-not-infusing^MDC||||R||20220517084831-
0700||||20220517084831-0700
OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-primary~^pump-source-info-
flush||||R||20220517084831-0700||||20220517084831-0700
OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-primary||||R||20220517084831-
0700||||20220517084831-0700
OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|0|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||
||R||20220517084831-0700||||20220517084831-0700
OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.2.0||||X||20220517084831-
0700||||20220517084831-0700
OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A||||R||20220517084831-
0700||||20220517084831-0700
OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.2.1|^pump-delivery-status-
transitioning^MDC||||R||20220517084831-0700||||20220517084831-0700

765 OBX|13|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.2.3|^pump-program-delivery-mode-continuous^MDC||||R||20220517084831-0700||||20220517084831-0700
OBX|15|NM|157784^MDC_FLOW_FLUID_PUMP^MDC|1.1.2.9|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||||R||20220517084831-0700||||20220517084831-0700
OBX|16|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.2.10|Clindamycin||||R||20220517084831-0700||||20220517084831-0700
770 OBX|17|SN|157760^MDC_CONC_DRUG^MDC|1.1.2.11|200.0000/2.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
OBX|18|ST|184515^MDC_DRUG_ID^MDC|1.1.2.15|Clindamycin||||R||20220517084831-0700||||20220517084831-0700
OBX|19|ST|184516^MDC_PUMP_DRUG_LIBRARY_CARE_AREA^MDC|1.1.2.16|Some Drug Library||||R||20220517084831-0700||||20220517084831-0700
775 OBX|20|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.2.31|2.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
OBX|21|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.2.32|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
OBX|22|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.2.33|2.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
780 OBX|23|NM|157924^MDC_RATE_DOSE^MDC|1.1.2.41|200.0000|263890^MDC_DIM_MILLI_G^MDC^^mg^UCUM||||R||20220517084831-0700||||20220517084831-0700
OBX|24|NM|157760^MDC_CONC_DRUG^MDC|1.1.2.111|100.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/mL^UCUM||||R||20220517084831-0700||||20220517084831-0700
785 OBX|25||70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700
OBX|26|NM|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700
OBX|27|NM|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-0700
OBX|28|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
790 OBX|29|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|30||0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700

795 OBX|34|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline||||R|||20220512135317-0700||||20220512135317-0700

OBX|35|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|Unknown||||R|||20220512135317-0700||||20220512135317-0700

OBX|36|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R|||20220512135317-0700||||20220512135317-0700

800 OBX|37|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R|||20220512135317-0700||||20220512135317-0700

OBX|38|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R|||20220512135317-0700||||20220512135317-0700

A.2.4 Flush Start Event (Immediately following Stop Event)

805 MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
 App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
 PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
 PID|||2392^^^HL7^PI||^U|||||

810 OBR|1|Unknown^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|999999^Medication_Unknown|||20220517084831-
 0700|||||MED0101|||||

OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0|||||X|||20220517084831-
 0700||||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700

OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0|||||X|||20220517084831-
 0700||||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700

815 OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197288^MDC_EVT_PUMP_DELIV_START^MDC||||R|||202205170
 84831-0700|||||20220517084831-0700

OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2|||||X|||20220517084831-0700|||||20220517084831-0700

OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0|||||X|||20220517084831-0700|||||20220517084831-0700

820 OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-infusing^MDC||||R|||20220517084831-
 0700|||||20220517084831-0700

OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-primary~^pump-source-info-
 flush||||R|||20220517084831-0700|||||20220517084831-0700

OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-flush||||R|||20220517084831-
 0700|||||20220517084831-0700

825 OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UC
 UM||||R|||20220517084831-0700|||||20220517084831-0700

OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.8.0|||||X|||20220512135317-
 0700|||||20220512135317-0700

830 OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A||||R|||20220517084831-
 0700|||||20220517084831-0700

OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.8.1|^pump-delivery-status-flushing^MDC||||R||20220512135317-0700||||20220512135317-0700

OBX|13|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.8.3|^pump-program-delivery-mode-continuous^MDC||||R||20220512135317-0700||||20220512135317-0700

835 OBX|14|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline||||R||20220512135317-0700||||20220512135317-0700

OBX|15|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|Unknown||||R||20220512135317-0700||||20220512135317-0700

OBX|16|NM|158084^MDC_VOL_FLUID_DELIV_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

840 OBX|17|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|18|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|19|70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700

845 OBX|20|NM|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700

OBX|21|ST|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-0700

OBX|22|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

850 OBX|23|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|24|0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700

OBX|25|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

855 OBX|26|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|27|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||2022
0512135317-0700||||20220512135317-0700

A.2.5 Flush Complete Event

860 MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
PID|||2392^^^HL7^PI||^U|||||||
865 OBR|1|MED0101^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|999999^Medication_Unknown|||20220517084831-
0700|||||MED0101|||||
OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0|||||X|||20220517084831-
0700||||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0|||||X|||20220517084831-
0700||||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
870 OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197290^MDC_EVT_PUMP_DELIV_STOP^MDC|||||R|||2022051708
4831-0700|||||20220517084831-0700
OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2|||||X|||20220517084831-0700|||||20220517084831-0700
OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0|||||X|||20220517084831-0700|||||20220517084831-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-not-infusing^MDC|||||R|||20220517084831-
875 0700|||||20220517084831-0700
OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-flush|||||R|||20220517084831-
0700|||||20220517084831-0700
OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-flush|||||R|||20220517084831-
0700|||||20220517084831-0700
880 OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|0|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||
||R|||20220517084831-0700|||||20220517084831-0700
OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC ^MDC|1.1.8.0|||||X|||20220512135317-
0700|||||20220512135317-0700
OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A|||||R|||20220517084831-
885 0700|||||20220517084831-0700

OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.8.1|^pump-delivery-status-not-delivering^MDC||||R||20220512135317-0700||||20220512135317-0700

OBX|13|CWE|158006^MDC_PUMP_NOT_DELIVERING_REASON^MDC|1.1.8.2|^pump-stopped-flush^MDC||||R||20220517092221-0700||||20220517092221-0700

890 OBX|14|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.8.3|^pump-program-delivery-mode-continuous^MDC||||R||20220512135317-0700||||20220512135317-0700

OBX|15|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline||||R||20220512135317-0700||||20220512135317-0700

OBX|16|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|Unknown||||R||20220512135317-0700||||20220512135317-0700

895 OBX|17|NM|158084^MDC_VOL_FLUID_DELIV_TOTAL^MDC|1.1.8.31|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|18|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

900 OBX|19|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|20||70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700

OBX|21|NM|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|syringe type||||R||20220512135317-0700||||20220512135317-0700

OBX|22|ST|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-0700

905 OBX|23|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|24|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|25||0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700

910 OBX|26|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

OBX|27|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||
|20220512135317-0700||||20220512135317-0700
OBX|28|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.4121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||2022
0512135317-0700||||20220512135317-0700

915

A.3 Auto Programming Flush After Infusion Complete

A.3.1 Flush Start Event

MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
 App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
 920 PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
 PID|||W.Yueh^^^HL7^PI||^U|||||||
 OBR|1|FLUSHIE2000^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|A10005^Normal Saline|||20220517084831-
 0700|||||
 925 OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0||||X|||20220517084831-
 0700||||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
 OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0||||X|||20220517084831-
 0700||||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
 OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197288^MDC_EVT_PUMP_DELIV_START^MDC||||R|||202205170
 84831-0700||||20220517084831-0700
 930 OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2||||X|||20220517084831-0700||||20220517084831-0700
 OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0||||X|||20220517084831-0700||||20220517084831-0700
 OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-infusing^MDC||||R|||20220517084831-
 0700||||20220517084831-0700
 935 OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-flush||||R|||20220517084831-
 0700||||20220517084831-0700
 OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-flush||||R|||20220517084831-
 0700||||20220517084831-0700
 OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UC
 UM||||R|||20220517084831-0700||||20220517084831-0700
 940 OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.8.0||||X|||20220512135317-
 0700||||20220512135317-0700
 OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A||||R|||20220517084831-
 0700||||20220517084831-0700

945 OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.8.1|^pump-delivery-status-flushing^MDC||||R||20220512135317-0700||||20220512135317-0700
OBX|13|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.8.3|^pump-program-delivery-mode-continuous^MDC||||R||20220512135317-0700||||20220512135317-0700
OBX|15|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline||||R||20220512135317-0700||||20220512135317-0700
950 OBX|16|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|A10005||||R||20220512135317-0700||||20220512135317-0700
OBX|17|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|18|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.5000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
955 OBX|19|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|20|70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700
OBX|21|ST|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700
OBX|22|ST|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-0700
960 OBX|23|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|2|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|24|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|2|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|25|0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700
965 OBX|27|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|28|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.5000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|29|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
970

A.3.2 Flush Complete Event

```

975 MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
PID||W.Yueh^^^HL7^PI||^U|||||
OBR|1|FLUSHIE2000^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|A10005^Normal Saline|||20220517084831-
0700|||||
980 OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0|||||X|||20220517084831-
0700|||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0|||||X|||20220517084831-
0700|||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197290^MDC_EVT_PUMP_DELIV_STOP^MDC|||||R|||2022051708
985 4831-0700|||||20220517084831-0700
OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2|||||X|||20220517084831-0700|||||20220517084831-0700
OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0|||||X|||20220517084831-0700|||||20220517084831-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-not-infusing^MDC|||||R|||20220517084831-
0700|||||20220517084831-0700
990 OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-flush|||||R|||20220517084831-
0700|||||20220517084831-0700
OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-flush|||||R|||20220517084831-
0700|||||20220517084831-0700
OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|0|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||
995 ||R|||20220517084831-0700|||||20220517084831-0700
OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.8.0|||||X|||20220512135317-
0700|||||20220512135317-0700
OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A|||||R|||20220517084831-
0700|||||20220517084831-0700

```

1000 OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.8.1|^pump-delivery-status-not-delivering^MDC||||R||20220512135317-0700||||20220512135317-0700
 OBX|13|CWE|158006^MDC_PUMP_NOT_DELIVERING_REASON^MDC|1.1.8.2|^pump-stopped-flush^MDC||||R||20220517092221-0700||||20220517092221-0700
 OBX|14|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.8.3|^pump-program-delivery-mode-continuous^MDC||||R||20220512135317-0700||||20220512135317-0700
 1005 OBX|15|NM|157784^MDC_FLOW_FLUID_PUMP^MDC|1.1.8.9|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||||R||20220517091144-0700||||20220517091144-0700
 OBX|16|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline||||R||20220512135317-0700||||20220512135317-0700
 1010 OBX|17|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|A10005||||R||20220512135317-0700||||20220512135317-0700
 OBX|18|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0.5000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
 OBX|19|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
 1015 OBX|20|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
 OBX|21|ST|70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700
 OBX|22|ST|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700
 OBX|23|NM|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-0700
 1020 OBX|24|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|2.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
 OBX|25|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|1.5000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
 1025 OBX|26||0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700
 OBX|27|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0.5000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
 OBX|28|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

1030 OBX|29|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||2022
0512135317-0700||||20220512135317-0700

A.4 Auto Programming Flush After Syringe Empty

A.4.1 Infusion Start Event

1035 MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
PID||W.Yueh^^^HL7^PI||^U|||||
1040 OBR|1|MED0101^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|Clindamycin^Clindamycin||20220517084831-
0700|||||
OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0||||X||20220517084831-
0700||||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0||||X||20220517084831-
0700||||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
1045 OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197288^MDC_EVT_PUMP_DELIV_START^MDC||||R||202205170
84831-0700||||20220517084831-0700
OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2||||X||20220517084831-0700||||20220517084831-0700
OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0||||X||20220517084831-0700||||20220517084831-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-infusing^MDC||||R||20220517084831-
1050 0700||||20220517084831-0700
OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-primary~^pump-source-info-
flush||||R||20220517084831-0700||||20220517084831-0700
OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-primary||||R||20220517084831-
0700||||20220517084831-0700
1055 OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UC
UM||||R||20220517084831-0700||||20220517084831-0700

OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.2.0||||X||20220517084831-0700||||20220517084831-0700

1060 OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A||||R||20220517084831-0700||||20220517084831-0700

OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.2.1|^pump-delivery-status-delivering^MDC||||R||20220517084831-0700||||20220517084831-0700

OBX|13|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.2.3|^pump-program-delivery-mode-continuous^MDC||||R||20220517084831-0700||||20220517084831-0700

1065 OBX|14|NM|157784^MDC_FLOW_FLUID_PUMP^MDC|1.1.2.9|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||||R||20220517084831-0700||||20220517084831-0700

OBX|15|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.2.10|Clindamycin||||R||20220517084831-0700||||20220517084831-0700

1070 OBX|16|SN|157760^MDC_CONC_DRUG^MDC|1.1.2.11|200.0000/2.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/mL^UCUM||||R||20220517084831-0700||||20220517084831-0700

OBX|17|ST|184515^MDC_DRUG_ID^MDC|1.1.2.15|Clindamycin||||R||20220517084831-0700||||20220517084831-0700

OBX|18|ST|184516^MDC_PUMP_DRUG_LIBRARY_CARE_AREA^MDC|1.1.2.16|Some Drug Library||||R||20220517084831-0700||||20220517084831-0700

1075 OBX|19|ST|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.2.31|1.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220517084831-0700||||20220517084831-0700

OBX|20|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.2.32|0.4879|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220517084831-0700||||20220517084831-0700

OBX|21|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.2.33|2.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220517084831-0700||||20220517084831-0700

1080 OBX|22|NM|157924^MDC_RATE_DOSE^^MDC|1.1.2.41|200.0000|263890^MDC_DIM_MILLI_G^MDC^^mg^UCUM||||R||20220517084831-0700||||20220517084831-0700

OBX|23|NM|157760^MDC_CONC_DRUG^MDC|1.1.2.111|100.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/mL^UCUM||||R||20220517084831-0700||||20220517084831-0700

OBX|24||70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700

1085 OBX|25|NM|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700

OBX|26|NM|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD|||||R||20220512135317-0700||||20220512135317-0700

OBX|27|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700

1090 OBX|28|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700

OBX|29|0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0|||||X||20220512135317-0700||||20220512135317-0700

OBX|33|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline|||||R||20220512135317-0700||||20220512135317-0700

1095 OBX|34|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|A10005|||||R||20220512135317-0700||||20220512135317-0700

OBX|35|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700

OBX|36|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700

1100 OBX|37|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM|||||R||20220512135317-0700||||20220512135317-0700

A.4.2 Infusion Complete Event

MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE

1105 PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO

PID||W.Yueh^^^HL7^PI||^U|||||

OBR|1|MED0101^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|Clindamycin^Clindamycin||20220517084831-0700|||||

OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0|||||X||20220517084831-0700||||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700

1110 OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0|||||X||20220517084831-0700||||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700

OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197290^MDC_EVT_PUMP_DELIV_COMP^MDC|||||R||20220517084831-0700||||20220517084831-0700

1115 OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2|||||X|||20220517084831-0700||||20220517084831-0700
OBX|5|70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0|||||X|||20220517084831-0700||||20220517084831-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-not-infusing^MDC|||||R|||20220517084831-
0700||||20220517084831-0700
OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-primary~^pump-source-info-
1120 flush|||||R|||20220517084831-0700||||20220517084831-0700
OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-primary|||||R|||20220517084831-
0700||||20220517084831-0700
OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|0|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||
||R|||20220517084831-0700||||20220517084831-0700
1125 OBX|10|70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.2.0|||||X|||20220517084831-
0700||||20220517084831-0700
OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A|||||R|||20220517084831-
0700||||20220517084831-0700
OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.2.1|^pump-delivery-status-
1130 transitioning^MDC|||||R|||20220517084831-0700||||20220517084831-0700
OBX|13|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.2.3|^pump-program-delivery-mode-
continuous^MDC|||||R|||20220517084831-0700||||20220517084831-0700
OBX|14|NM|157784^MDC_FLOW_FLUID_PUMP^MDC|1.1.2.9|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^U
CUM|||||R|||20220517084831-0700||||20220517084831-0700
1135 OBX|15|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.2.10|Clindamycin|||||R|||20220517084831-0700||||20220517084831-
0700
OBX|16|SN|157760^MDC_CONC_DRUG^MDC|1.1.2.11|200.0000/2.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/m
L^UCUM|||||R|||20220517084831-0700||||20220517084831-0700
OBX|17|ST|184515^MDC_DRUG_ID^MDC|1.1.2.15|Clindamycin|||||R|||20220517084831-0700||||20220517084831-0700
1140 OBX|18|ST|184516^MDC_PUMP_DRUG_LIBRARY_CARE_AREA^MDC|1.1.2.16|Some Drug Library|||||R|||20220517084831-
0700||||20220517084831-0700
OBX|19|ST|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.2.31|2.0000|263762^MDC_DIM_MILLI_L^MDC^
^mL^UCUM|||||R|||20220517084831-0700||||20220517084831-0700

1145 OBX|20|ST|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.2.32|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||
20220517084831-0700||||20220517084831-0700
OBX|21|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.2.33|2.0000|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||2022
0517084831-0700||||20220517084831-0700
OBX|22|NM|157924^MDC_RATE_DOSE^^MDC|1.1.2.41|200.0000|263890^MDC_DIM_MILLI_G^MDC^^mg^UCUM||||R||2022
0517084831-0700||||20220517084831-0700
1150 OBX|23|NM|157760^MDC_CONC_DRUG^MDC|1.1.2.111|100.0000|264306^MDC_DIM_MILLI_G_PER_ML^MDC^^mg/mL^U
CUM||||R||20220517084831-0700||||20220517084831-0700
OBX|24||70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700
OBX|25|NM|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700
OBX|26|NM|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-
1155 0700
OBX|27|NM|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135
317-0700||||20220512135317-0700
OBX|28|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||2
0220512135317-0700||||20220512135317-0700
1160 OBX|29||0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700
OBX|33|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline||||R||20220512135317-0700||||20220512135317-
0700
OBX|34|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|A10005||||R||20220512135317-0700||||20220512135317-0700
OBX|35|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL
1165 ^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|36|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCU
M||||R||20220512135317-0700||||20220512135317-0700
OBX|37|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||2022
0512135317-0700||||20220512135317-0700
1170

A.4.3 Flush Start Event

MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
1175 PID||W.Yueh^^^HL7^PI||^U|||||||
OBR|1|FLUSHIE2000^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|A10005^Normal Saline||20220517084831-
0700|||||MED0101|||||
OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0||||X||20220517084831-
0700||||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
1180 OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0||||X||20220517084831-
0700||||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197288^MDC_EVT_PUMP_DELIV_START^MDC||||R||202205170
84831-0700||||20220517084831-0700
OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2||||X||20220517084831-0700||||20220517084831-0700
1185 OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0||||X||20220517084831-0700||||20220517084831-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-infusing^MDC||||R||20220517084831-
0700||||20220517084831-0700
OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-flush||||R||20220517084831-
0700||||20220517084831-0700
1190 OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-flush||||R||20220517084831-
0700||||20220517084831-0700
OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|1.3333|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UC
UM||||R||20220517084831-0700||||20220517084831-0700
OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.8.0||||X||20220512135317-
1195 0700||||20220512135317-0700
OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A||||R||20220517084831-
0700||||20220517084831-0700
OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.8.1|^pump-delivery-status-
flushing^MDC||||R||20220512135317-0700||||20220512135317-0700

1200 OBX|13|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.8.3|^pump-program-delivery-mode-continuous^MDC||||R||20220512135317-0700||||20220512135317-0700
OBX|14|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline||||R||20220512135317-0700||||20220512135317-0700
OBX|15|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|A10005||||R||20220512135317-0700||||20220512135317-0700
1205 OBX|16|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|17|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|18|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
1210 OBX|19||70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700
OBX|20|NM|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700
OBX|21|ST|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-0700
OBX|22|ST|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
1215 OBX|23|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|24||0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700
OBX|25|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
1220 OBX|26|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|27|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
1225

A.4.4 Flush Complete Event

1230 MSH|^~\&|DEVICE_PUMPVENDOR^00D07500001^EUI-64|PUMPVENDOR^00D07500001^EUI-64|Receiving
App|HOSTFACILITY|20220517084831-0700||ORU^R42^ORU_R01|10e8590d-298c-4047-bb13-1d923bd001c3|P|2.6||AL|NE||||IHE
PCD-10^IHE PCD^1.3.6.1.4.1.19376.1.6.4.10^ISO
PID|||W.Yueh^^^HL7^PI||^U|||||||
OBR|1|FLUSHIE2000^HL7^00D0750000000154^EUI-64|744^SyrPump^744^EUI-64|A10005^Normal Saline||20220517084831-
0700|||||MED0101|||||
OBX|1||69985^MDC_DEV_PUMP_INFUS_MDS^MDC|1.0.0.0|||||X||20220517084831-
0700||||M8110S1298100^SyrPump^00D075156BB466B4^EUI-64|20220517084831-0700
1235 OBX|2||70054^MDC_DEV_PUMP_INFUS_SYRINGE_VMD^MDC|1.1.0.0|||||X||20220517084831-
0700||||M8110S54321^SyrPump^00D0750036E62C71^EUI-64|20220517084831-0700
OBX|3|CWE|68487^MDC_ATTR_EVT_COND^MDC|1.1.0.1|197290^MDC_EVT_PUMP_DELIV_STOP^MDC|||||R||2022051708
4831-0700||||20220517084831-0700
OBX|4|ST|68488^MDC_ATTR_EVT_SOURCE^MDC|1.1.0.2|||||X||20220517084831-0700||||20220517084831-0700
1240 OBX|5||70067^MDC_DEV_PUMP_DELIVERY_INFO^MDC|1.1.1.0|||||X||20220517084831-0700||||20220517084831-0700
OBX|6|CWE|184519^MDC_PUMP_INFUSING_STATUS^MDC|1.1.1.1|^pump-status-not-infusing^MDC|||||R||20220517084831-
0700||||20220517084831-0700
OBX|7|CWE|158015^MDC_PUMP_ENABLED_SOURCES^MDC|1.1.1.2|^pump-source-info-flush|||||R||20220517084831-
0700||||20220517084831-0700
1245 OBX|8|CWE|158016^MDC_PUMP_ACTIVE_SOURCES^MDC|1.1.1.3|^pump-source-info-flush|||||R||20220517084831-
0700||||20220517084831-0700
OBX|9|NM|158014^MDC_FLOW_FLUID_PUMP^MDC|1.1.1.2|0|265266^MDC_DIM_MILLI_L_PER_HR^MDC^^mL/h^UCUM||
||R||20220517084831-0700||||20220517084831-0700
OBX|10||70071^MDC_DEV_PUMP_INFUSATE_SOURCE_PRIMARY_CHAN^MDC|1.1.8.0|||||X||20220512135317-
1250 0700||||20220512135317-0700
OBX|11|ST|158012^MDC_PUMP_SOURCE_CHANNEL_LABEL^MDC|1.1.1.52|A|||||R||20220517084831-
0700||||20220517084831-0700
OBX|12|CWE|158005^MDC_PUMP_CURRENT_DELIVERY_STATUS^MDC|1.1.8.1|^pump-delivery-status-not-
delivering^MDC|||||R||20220512135317-0700||||20220512135317-0700

1255 OBX|12|CWE|158006^MDC_PUMP_NOT_DELIVERING_REASON^MDC|1.1.8.2|^pump-stopped-flush^MDC||||R||20220517092221-0700||||20220517092221-0700
OBX|13|CWE|158008^MDC_PUMP_PROGRAM_DELIVERY_MODE^MDC|1.1.8.3|^pump-program-delivery-mode-continuous^MDC||||R||20220512135317-0700||||20220512135317-0700
OBX|14|ST|184514^MDC_DRUG_NAME_LABEL^MDC|1.1.8.10|Normal Saline||||R||20220512135317-0700||||20220512135317-0700
1260 OBX|15|ST|184515^MDC_DRUG_ID^MDC|1.1.8.15|A10005||||R||20220512135317-0700||||20220512135317-0700
OBX|16|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|17|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
1265 OBX|18|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|19||70091^MDC_DEV_PUMP_SYRINGE_INFO^MDC|1.1.9.0||||X||20220512135317-0700||||20220512135317-0700
OBX|20|NM|184476^MDC_SYRINGE_TYPE^MDC|1.1.9.1|BD-10||||R||20220512135317-0700||||20220512135317-0700
1270 OBX|21|ST|184488^MDC_SYRINGE_MANUFACTURER^MDC|1.1.9.2|BD||||R||20220512135317-0700||||20220512135317-0700
OBX|22|ST|157880^MDC_VOL_SYRINGE^MDC|1.1.9.3|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|23|NM|157984^MDC_VOL_SYRINGE_ACTUAL^MDC|1.1.9.4|1|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
1275 OBX|24||0^MDCX_DEV_PUMP_FLUSH_INFO^MDC|1.1.10.0||||X||20220512135317-0700||||20220512135317-0700
OBX|25|NM|158084^MDC_VOL_FLUID_DELIV_METHOD_TOTAL^MDC|1.1.8.31|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
OBX|26|NM|157872^MDC_VOL_FLUID_TBI_REMAIN^MDC|1.1.8.32|0|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700
1280 OBX|27|NM|157884^MDC_VOL_FLUID_TBI^MDC|1.1.8.33|0.5121|263762^MDC_DIM_MILLI_L^MDC^^mL^UCUM||||R||20220512135317-0700||||20220512135317-0700

Appendix B – Auto-Programming Request (APR) in HL7 format

B.1 FLUSH APR PRN

```

1285      MSH|^~\&|EMR|EMR|DEVICE_PUMPVENDOR|PUMPVENDOR|20220520130034-
        0600||RGV^O15^RGV_O15|10|P|2.6||AL|AL||||IHE_PCD_RGV_O15^IHE PCD^1.3.6.1.4.1.19376.1.6.1.3.1^ISO|
        PID|||W.Yueh^AE^^PI^PI||W^Yueh^^^^^L||19710101|M||15 N
        SAGUARO^^TUCSON^AZ^85701^USA^^^PIMA|PIMA||||AccountId^B^||||||N|||||
        ORC|RE|FLUSHIE2000|||||20220520130034||||||RatchedRN^R^RN^^
1290      RXG|1||A10005^Normal Saline|0.5000||263762^MDC_DIM_MILLI_L^MDC|||||||
        RXR|^IV^HL70162||^SYR^HL70164|^IV^HL70165
        OBX|1||69986^MDC_DEV_PUMP_INFUS_VMD^MDC|||||||M8110S12382855^^M8110S12382855^
        OBX|2||158038^MDC_INFUS_ORDER_TYPE^MDC||continuous|||||||^^^

```

B.2 FLUSH APR

```

1295      MSH|^~\&|EMR|EMR|DEVICE_PUMPVENDOR |PUMPVENDOR|20220520130034-
        0600||RGV^O15^RGV_O15|10|P|2.6||AL|AL||||IHE_PCD_RGV_O15^IHE PCD^1.3.6.1.4.1.19376.1.6.1.3.1^ISO|
        PID|||W.Yueh^AE^^PI^PIE||W^Yueh^^^^^L||19710101|M||15 N
        SAGUARO^^TUCSON^AZ^85701^USA^^^PIMA|PIMA||||AccountId^B^||||||N|||||
        ORC|RE|FLUSHIE2000|||||20220520130034||||||RatchedRN ^R^RN^^
1300      RXG|1||A10005^Normal Saline |0.5000||263762^MDC_DIM_MILLI_L^MDC|||||||
        RXR|^IV^HL70162||^SYR^HL70164|^IV^HL70165
        OBX|1||69986^MDC_DEV_PUMP_INFUS_VMD^MDC|||||||M8110S12382855^^M8110S12382855^
        OBX|2||158038^MDC_INFUS_ORDER_TYPE^MDC||continuous|||||||^^^

```

1305 **Glossary**

The complete IHE Glossary is available [here](#).

Term	Definition
alarm	A clinical alarm is an indication from a system or device, that when activated, indicates a condition requiring urgent clinical assessment and possible intervention.
Auto Program	A pump program in which some or all settings are received from another system such as an eMAR or BCMA system.
event	An occurrence about which it is desired to communicate information between devices and information systems.
Health Level 7	Founded in 1987, Health Level Seven International (HL7) is a not-for-profit, ANSI-accredited standards developing organization dedicated to providing a comprehensive framework and related standards for the exchange, integration, sharing, and retrieval of electronic health information that supports clinical practice and the management, delivery and evaluation of health services.
PatientID	(When used in the context of ATNA) A free text that holds the system-internal patient identifier being unique within that system domain. The patient identifier domain is that assigned to the system that generated the audit event record. The patient information in audit event records corresponds to the information available to identify a patient at the time the audit record was generated and does not reflect later updates (e.g., patient reconciliation)