

Integrating the Healthcare Enterprise



5

**IHE Radiation Oncology (RO)
Technical Framework**

10

**Volume 3
IHE RO TF-3
Content Modules**

15

20

**Revision 2.0 - Final Text
April 7, 2020**

25

Please verify you have the most recent version of this document, which is published [here](#).

CONTENTS

30	1	Introduction	9
	1.1	Introduction to IHE	9
	1.2	Intended Audience	9
	1.3	Overview of Technical Framework Volume 3	9
	1.4	Comment Process.....	10
35	1.5	Copyright Licenses	10
	1.5.1	Copyright of Base Standards	10
	1.5.1.1	DICOM (Digital Imaging and Communications in Medicine)	11
	1.5.1.2	HL7 (Health Level Seven).....	11
	1.5.1.3	LOINC (Logical Observation Identifiers Names and Codes)	11
40	1.5.1.4	SNOMED CT (Systematized Nomenclature of Medicine -- Clinical Terms)....	11
	1.6	Trademark	11
	1.7	Disclaimer Regarding Patent Rights.....	11
	1.8	History of Document Changes.....	12
	2	Conventions	13
45	2.1	Content Module Modeling and Profiling Conventions.....	13
	2.2	Additional Standards Profiling Conventions	13
	3	Content Modules Overview and Terminology.....	14
	4	Reserved for domain specific content	15
	5	IHE Namespaces, Concept Domains, and Vocabularies	15
50	5.1	IHE Radiation Oncology Namespaces.....	15
	5.2	IHE Radiation Oncology Concept Domains.....	15
	5.3	IHE Radiation Oncology Format Codes and Vocabularies	15
	5.3.1	IHE Format Codes	15
	5.3.2	IHEActCode Vocabulary.....	15
55	5.3.3	IHERoleCode Vocabulary	15
	6	IHE RO HL7 V3 CDA Content Modules	16
	7	IHE RO DICOM Content Modules	17
	7.1	Conventions	17
	7.1.1	Scope of Requirements	17
60	7.1.2	Requirements Definitions	17
	7.1.3	Display Requirements.....	18
	7.2	General Definitions.....	18
	7.2.1	Character Sets	18
	7.2.1.1	Support of Character Sets other than ISO-IR 100	18
65	7.2.2	Propagation of Common Patient Information.....	18
	7.2.3	Study Handling.....	19
	7.2.4	Frame of Reference Handling.....	19
	7.3	IOD Definitions	19
	7.3.1	Prescription IODs	19
70	7.3.2	Plan IODs.....	20

	7.3.2.1	Technique Specific RT Plan IODs	20
	7.3.2.1.1	RT Plan IOD for Photon External Beam in Planning State	20
	7.3.2.1.1.1	Referenced Standards	20
	7.3.2.1.1.2	IOD Definition.....	20
75	7.3.2.1.2	RT Plan IOD for Photon External Beam in Delivery State	21
	7.3.2.2	RT Plan IOD for General Use	21
	7.3.2.2.1	RT Plan IOD from Dosimetric Planning	21
	7.3.2.2.1.1	Referenced Standards.....	21
	7.3.2.2.1.2	IOD Definition.....	22
80	7.3.2.2.2	RT Plan IOD for Dose Composition.....	23
	7.3.2.2.3	RT Plan IOD for Consistent Dose Tracking.....	23
	7.3.2.2.4	RT Ion Plan IOD from Dosimetric Planning	23
	7.3.2.2.4.1	Referenced Standards	23
	7.3.2.2.4.2	IOD Definition.....	24
85	7.3.2.2.5	RT Plan IOD from Geometric Planning	25
	7.3.2.2.5.1	Referenced Standards	25
	7.3.2.2.5.2	Definition.....	26
	7.3.3	Image IOD	27
	7.3.3.1	RT Image.....	27
90	7.3.3.2	CT Image.....	27
	7.3.3.2.1	CT Image in Planning State	27
	7.3.3.2.2	CT Image in Delivery State	27
	7.3.3.2.3	CT Image for General Use.....	27
	7.3.3.2.3.1	Referenced Standards	27
95	7.3.3.2.3.2	IOD Definition.....	27
	7.3.4	RT Structure Set IOD	28
	7.3.4.1	RT Structure Set for General Use.....	28
	7.3.4.1.1	RT Structure Set for Basic Interoperability	28
	7.3.4.1.1.1	Referenced Standards	28
100	7.3.4.1.1.2	IOD Definition.....	29
	7.3.4.1.2	RT Structure Set Multi Modality Content	30
	7.3.4.1.2.1	Referenced Standards	30
	7.3.4.1.2.2	IOD Definition.....	30
	7.3.5	RT Dose IOD.....	31
105	7.3.5.1	RT Dose IOD for General Use	31
	7.3.5.1.1	RT Dose from Dosimetric Planning	31
	7.3.5.1.1.1	Referenced Standards	31
	7.3.5.1.1.2	IOD Definition.....	31
	7.3.6	Treatment Record	32
110	7.3.7	Reporting IOD	32
	7.3.8	ROI Dictionary IOD	32
	7.3.9	Workflow IOD.....	32
	7.3.10	Spatial Registration IOD	32
	7.3.10.1	Spatial Registration IOD for General Use	32

115	7.3.10.1.1 Spatial Registration IOD Base Content	32
	7.3.10.1.1.1 Referenced Standards	32
	7.3.10.1.1.2 IOD Definition	33
	7.4 Module Definitions	33
	7.4.1 General Modules	33
120	7.4.1.1 Patient Module	33
	7.4.1.1.1 Patient Module Base Content	33
	7.4.1.1.1.1 Referenced Standards	33
	7.4.1.1.1.2 Module Definition	33
	7.4.1.2 Study Module	34
125	7.4.1.2.1 General Study Module Base Content	34
	7.4.1.2.1.1 Referenced Standards	34
	7.4.1.2.1.2 Module Definition	34
	7.4.1.3 General Series Module	35
	7.4.1.3.1 General Series Module Base Content	35
130	7.4.1.3.1.1 Referenced Standards	35
	7.4.1.3.1.2 Module Definition	35
	7.4.1.3.2 General Series Module Feet First	35
	7.4.1.3.2.1 Referenced Standards	35
	7.4.1.3.2.2 Module Definition	35
135	7.4.1.3.3 General Series Module Decubitus	35
	7.4.1.3.3.1 Referenced Standards	35
	7.4.1.3.3.2 Module Definition	36
	7.4.1.4 RT Series Module	36
	7.4.1.4.1 RT Series Module Base Content	36
140	7.4.1.4.1.1 Referenced Standards	36
	7.4.1.4.1.2 Module Definition	36
	7.4.1.5 Equipment Module	36
	7.4.1.5.1 General Equipment Module Base Content	36
	7.4.1.5.1.1 Referenced Standards	36
145	7.4.1.5.1.2 Module Definition	36
	7.4.1.6 SOP Common Module	37
	7.4.1.6.1 SOP Common Module Base Content	37
	7.4.1.6.1.1 Referenced Standards	37
	7.4.1.6.1.2 Module Definition	37
150	7.4.1.7 Frame of Reference Module	37
	7.4.1.7.1 Frame of Reference Module Base Content	37
	7.4.1.7.1.1 Referenced Standards	37
	7.4.1.7.1.2 Module Definition	38
	7.4.1.8 General Image Module	38
155	7.4.1.8.1 General Image Module Base Content	38
	7.4.1.8.1.1 Referenced Standards	38
	7.4.1.8.1.2 Module Definition	38
	7.4.2 Workflow-Related Modules	38

	7.4.3 General Plan-Related Modules	38
160	7.4.3.1 General Plan Module	38
	7.4.3.1.1 General Plan Module Base Content	38
	7.4.3.1.1.1 Referenced Standards	38
	7.4.3.1.1.2 Module Definition	38
	7.4.3.2 RT Prescription Module	39
165	7.4.3.2.1 RT Prescription Module Base Content	39
	7.4.3.2.1.1 Referenced Standards	39
	7.4.3.2.1.2 Module Definition	39
	7.4.3.3 RT Fraction Scheme Module.....	39
	7.4.3.3.1 RT Fraction Scheme Module for Consistent Dose	39
170	7.4.3.3.2 RT Fraction Scheme Module for Delivery	39
	7.4.3.3.2.1 Referenced Standards	39
	7.4.3.3.2.2 Module Definition	39
	7.4.3.3.3 RT Fraction Scheme Module for Brachy.....	40
	7.4.3.3.4 RT Fraction Scheme Module Base Content.....	40
175	7.4.3.3.4.1 Referenced Standards	40
	7.4.3.3.4.2 Module Definition	40
	7.4.3.4 RT Patient Setup Module	40
	7.4.3.4.1 RT Patient Setup Module Base Content	40
	7.4.3.4.1.1 Referenced Standards	40
180	7.4.3.4.1.2 Module Definition	41
	7.4.3.4.2 RT Patient Setup Module Feet First.....	41
	7.4.3.4.2.1 Referenced Standards	41
	7.4.3.4.2.2 Module Definition	41
	7.4.3.4.3 RT Patient Setup Module Reoriented	42
185	7.4.3.4.3.1 Referenced Standards	42
	7.4.3.4.3.2 Module Definition	42
	7.4.3.4.4 RT Patient Setup Module Decubitus.....	42
	7.4.3.4.4.1 Referenced Standards	42
	7.4.3.4.4.2 Module Definition	42
190	7.4.4 Plan-Related Modules in Planning	43
	7.4.4.1 Specific RT Beam Type Specifications.....	43
	7.4.4.1.1 RT Beams Module for Basic Static Beam	43
	7.4.4.1.1.1 Referenced Standards	43
	7.4.4.1.1.2 Module Definition	43
195	7.4.4.1.2 RT Beams Module for Basic Static MLC Beam	45
	7.4.4.1.2.1 Referenced Standards	45
	7.4.4.1.2.2 Module Definition	45
	7.4.4.1.3 RT Beams Module for Arc Beam	47
	7.4.4.1.3.1 Referenced Standards	47
200	7.4.4.1.3.2 Module Definition	47
	7.4.4.1.4 RT Beams Module for MLC Fixed Aperture Arc Beam	49
	7.4.4.1.4.1 Referenced Standards	49

	7.4.4.1.4.2	Module Definition	49
205	7.4.4.1.5	RT Beams Module for MLC Variable Aperture Arc Beam	51
	7.4.4.1.5.1	Referenced Standards	51
	7.4.4.1.5.2	Module Definition	51
	7.4.4.1.6	RT Beams Module for Hard Wedge Beam.....	52
	7.4.4.1.6.1	Referenced Standards	52
210	7.4.4.1.6.2	Module Definition	53
	7.4.4.1.7	RT Beams Module for Virtual Wedge Beam	55
	7.4.4.1.7.1	Referenced Standards	55
	7.4.4.1.7.2	Module Definition	55
	7.4.4.1.8	RT Beams Module for Motorized Wedge Beam	57
215	7.4.4.1.8.1	Referenced Standards	57
	7.4.4.1.8.2	Module Definition	57
	7.4.4.1.9	RT Beams Module for Basic Static Electron Beam.....	59
	7.4.4.1.9.1	Referenced Standards	59
	7.4.4.1.9.2	Module Definition	60
	7.4.4.1.10	RT Beams Module for Step & Shoot Beam.....	62
220	7.4.4.1.10.1	Referenced Standards	62
	7.4.4.1.10.2	Module Definition	62
	7.4.4.1.11	RT Beams Module for Sliding Window Beam.....	64
	7.4.4.1.11.1	Referenced Standards	64
	7.4.4.1.11.2	Module Definition	64
225	7.4.4.1.12	RT Beams Module for IMAT/VMAT Beam	66
	7.4.4.1.12.1	Referenced Standards	66
	7.4.4.1.12.2	Module Definition	66
	7.4.4.1.13	RT Beams Module for Photon Applicator Beam.....	68
230	7.4.4.1.13.1	Referenced Standards	68
	7.4.4.1.13.2	Module Definition	68
	7.4.4.1.14	RT Beams Module for Photon Applicator Arc Beam.....	70
	7.4.4.1.14.1	Referenced Standards	70
	7.4.4.1.14.2	Module Definition	70
235	7.4.4.2	General Beam Attribute Specifications	72
	7.4.4.2.1	Control Point Fixed Attribute List Base Content.....	72
	7.4.4.2.1.1	Referenced Standards	72
	7.4.4.2.1.2	Required Attributes.....	72
	7.4.4.3	Beam Option Specifications	73
240	7.4.4.3.1	Bolus Beam Modifier Base Content	73
	7.4.4.3.1.1	Referenced Standards	73
	7.4.4.3.1.2	Required Attributes.....	73
	7.4.4.3.2	Block Beam Modifier Base Content	74
	7.4.4.3.2.1	Referenced Standards	74
	7.4.4.3.2.2	Required Attributes.....	74
245	7.4.4.3.3	Compensator Beam Modifier Base Content	74
	7.4.4.3.3.1	Referenced Standards	74

	7.4.4.3.3.2 Required Attributes.....	75
	7.4.4.3.4 Hard Wedge Beam Modifier Base Content	75
250	7.4.4.3.4.1 Referenced Standards	75
	7.4.4.3.4.2 Required Attributes.....	75
	7.4.4.4 Other RT Beam Modules.....	76
	7.4.4.4.1 RT Beams Module for Geometric Planner	76
	7.4.4.4.1.1 Referenced Standards	76
	7.4.4.4.1.2 Module Definition	76
255	7.4.5 Plan-Related Modules in Delivery.....	78
	7.4.6 Image-Related Modules in Planning	78
	7.4.6.1 RT Image Module.....	78
	7.4.6.2 Image Plane Module.....	78
	7.4.6.2.1 Image Plane Base Content	78
260	7.4.6.2.1.1 Referenced Standards	78
	7.4.6.2.1.2 Module Definition	79
	7.4.6.2.2 Image Plane Decubitus	79
	7.4.6.2.2.1 Referenced Standards	79
	7.4.6.2.2.2 Module Definition	79
265	7.4.7 Image-Related Modules in Delivery.....	79
	7.4.8 Segment-Related Modules.....	79
	7.4.8.1 RT ROI Observation Module.....	79
	7.4.8.1.1 RT ROI Observation Module Base Content	79
	7.4.8.1.1.1 Referenced Standards	79
270	7.4.8.1.1.2 Module Definition	80
	7.4.8.2 RT ROI Contour Module.....	81
	7.4.8.2.1 RT ROI Contour Module Base Content.....	81
	7.4.8.2.1.1 Referenced Standards	81
	7.4.8.2.1.2 Module Definition	81
275	7.4.8.2.2 RT ROI Contour Module Off-slice.....	82
	7.4.8.2.2.1 Referenced Standards	82
	7.4.8.2.2.2 Module Definition	82
	7.4.8.3 RT Structure Set Module.....	83
	7.4.8.3.1 Structure Set Module Base Content.....	83
280	7.4.8.3.1.1 Referenced Standards	83
	7.4.8.3.1.2 Module Definition	84
	7.4.8.3.2 Structure Set in Multi Modality	85
	7.4.8.3.2.1 Referenced Standards	85
	7.4.8.3.2.2 Module Definition	85
285	7.4.9 Segment Modules in Delivery	86
	7.4.10 Registration Modules in Planning	86
	7.4.10.1 Registration Module.....	86
	7.4.10.1.1 Registration Module Base Content	86
	7.4.10.1.1.1 Referenced Standards	86
290	7.4.10.1.1.2 Module Definition	86

	7.4.11	Treatment Records.....	87
	7.4.12	Prescription-Related Modules in Planning.....	87
	7.4.13	Dose-Related Modules	87
	7.4.13.1	Image Plane Module	87
295	7.4.13.1.1	Image Plane Base Content	87
	7.4.13.1.1.1	Referenced Standards	87
	7.4.13.1.1.2	Module Definition	87
	7.4.13.2	Multi-Frame Module.....	88
	7.4.13.2.1	Multi-Frame Module Base Content	88
300	7.4.13.2.1.1	Referenced Standards	88
	7.4.13.2.1.2	Module Definition	88
	7.4.13.3	RT Dose Module.....	88
	7.4.13.3.1	RT Dose Module Base Content	88
	7.4.13.3.1.1	Referenced Standards	88
305	7.4.13.3.1.2	Module Definition	88
	7.4.13.4	RT DVH Module	89
	7.4.13.4.1	RT DVH Module Base Content.....	89
	7.4.13.4.1.1	Referenced Standard.....	89
	7.4.13.4.1.2	Module Definition	89
310	8	IHE RO Content Modules (type other than CDA or DICOM)	90
		Appendices.....	91
		Glossary	92

315 **1 Introduction**

This document, Volume 3 of the IHE Radiation Oncology (RO) Technical Framework, defines content modules used in the IHE Radiation Oncology profiles.

1.1 Introduction to IHE

320 Integrating the Healthcare Enterprise (IHE) is an international initiative to promote the use of standards to achieve interoperability among health information technology (HIT) systems and effective use of electronic health records (EHRs). IHE provides a forum for care providers, HIT experts and other stakeholders in several clinical and operational domains to reach consensus on standards-based solutions to critical interoperability issues.

325 The primary output of IHE is system implementation guides, called IHE Profiles. IHE publishes each profile through a well-defined process of public review and trial implementation and gathers profiles that have reached final text status into an IHE Technical Framework, of which this volume is a part.

330 For more general information regarding IHE, refer to www.ihe.net. It is strongly recommended that, prior to reading this volume, the reader familiarizes themselves with the concepts defined in the [IHE Technical Frameworks General Introduction](#).

1.2 Intended Audience

The intended audience of IHE Technical Frameworks Volume 3 is:

- IT departments of healthcare institutions
- Technical staff of vendors participating in the IHE initiative
- 335 • Experts involved in standards development

1.3 Overview of Technical Framework Volume 3

Volume 3 is comprised of several distinct sections:

- Section 1 provides background and reference material.
- Section 2 presents the conventions used in this volume to define the content modules.
- 340 • Section 3 provides an overview of Content Modules and the terminology used.
- Section 4 is reserved for domain unique Content Module specifications.
- Section 5 lists the namespaces and identifiers defined or referenced and the vocabularies defined or referenced herein.
- Section 6 defines <this domain's> HL7 V3 CDA Content Modules in detail.
- 345 • Section 7 defines <this domain's> DICOM content modules.
- Section 8 defines other types of content modules.

The appendices in Volume 3 provide clarification of technical details of the IHE data model and transactions. A glossary of terms and acronyms used in the IHE Technical Framework, including those from relevant standards, is provided in [Appendix D](#) to the *IHE Technical Frameworks General Introduction*. Due to the length of the document, some domains may divide Volume 3 into smaller volumes labeled 3a, 3b, etc. In this case, the Volume 3 appendices are gathered in Volume 3x. Code and message samples may also be stored on the IHE ftp server. In this case, explicit links to the ftp server will be provided in the transaction text.

1.4 Comment Process

IHE International and AAPM welcomes comments on this document and the IHE-RO initiative. They should be submitted at [http://www.ihe.net/Radiation Oncology Public Comments](http://www.ihe.net/Radiation_Oncology_Public_Comments) or to:

Jill I. Moton, MBA
Program Manager
American Association of Physicists in Medicine (AAPM)
1631 Prince Street
Alexandria, VA 22314
jill@aapm.org

1.5 Copyright Licenses

IHE International hereby grants to each Member Organization, and to any other user of these documents, an irrevocable, worldwide, perpetual, royalty-free, nontransferable, nonexclusive, non-sublicensable license under its copyrights in any IHE profiles and Technical Framework documents, as well as any additional copyrighted materials that will be owned by IHE International and will be made available for use by Member Organizations, to reproduce and distribute (in any and all print, electronic or other means of reproduction, storage or transmission) such IHE Technical Documents.

The licenses covered by this Copyright License are only to those copyrights owned or controlled by IHE International itself. If parts of the Technical Framework are included in products that also include materials owned or controlled by other parties, licenses to use those products are beyond the scope of this IHE document and would have to be obtained from that other party.

1.5.1 Copyright of Base Standards

IHE technical documents refer to and make use of a number of standards developed and published by several standards development organizations. All rights for their respective base standards are reserved by these organizations. This agreement does not supersede any copyright provisions applicable to such base standards. Copyright license information for frequently referenced base standards is provided below.

1.5.1.1 DICOM (Digital Imaging and Communications in Medicine)

DICOM[®] is the registered trademark of the National Electrical Manufacturers Association for its standards publications relating to digital communications of medical information.

1.5.1.2 HL7 (Health Level Seven)

385 HL7[®], Health Level Seven[®], CDA[®], FHIR[®], and the FHIR [FLAME DESIGN][®] are registered trademarks of Health Level Seven International.

Health Level Seven, Inc. has granted permission to IHE to reproduce tables from the HL7 standard. The HL7 tables in this document are copyrighted by Health Level Seven, Inc. All rights reserved. Material drawn from these documents is credited where used.

390 1.5.1.3 LOINC (Logical Observation Identifiers Names and Codes)

LOINC[®] is registered United States trademarks of Regenstrief Institute, Inc.

1.5.1.4 SNOMED CT (Systematized Nomenclature of Medicine -- Clinical Terms)

395 Some IHE Profiles incorporate SNOMED[®] CT, which is used by permission of the International Health Terminology Standards Development Organisation. SNOMED CT[®] was originally created by the College of American Pathologists. SNOMED CT is a registered trademark of the International Health Terminology Standards Development Organisation, all rights reserved.

1.6 Trademark

400 IHE[®] and the IHE logo are trademarks of the Healthcare Information Management Systems Society in the United States and trademarks of IHE Europe in the European Community. They may only be used with the written consent of the IHE International Board Operations Committee, which may be given to a Member Organization in broad terms for any use that is consistent with the IHE mission and operating principles.

1.7 Disclaimer Regarding Patent Rights

405 Attention is called to the possibility that implementation of the specifications in this document may require use of subject matter covered by patent rights. By publication of this document, no position is taken with respect to the existence or validity of any patent rights in connection therewith. IHE International is not responsible for identifying Necessary Patent Claims for which a license may be required, for conducting inquiries into the legal validity or scope of Patents Claims or determining whether any licensing terms or conditions provided in connection with
410 submission of a Letter of Assurance, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of the specifications in this document are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information about the IHE International patent disclosure process including links to forms for making disclosures is available at

415 http://www.ihe.net/Patent_Disclosure_Process. Please address questions about the patent disclosure process to the secretary of the IHE International Board: secretary@ihe.net.

1.8 History of Document Changes

This section provides a brief summary of changes and additions to this document.

Date	Document Revision	Change Summary
2020-04-07	2.0	Added initial content for Basic RT Objects – II, Treatment Planning Plan Content and Multi-Modality Registration 2018 profiles.

420

2 Conventions

425 This document has adopted the following conventions for representing the framework concepts and specifying how the standards upon which the IHE Technical Framework is based shall be applied.

2.1 Content Module Modeling and Profiling Conventions

430 In order to maintain consistent documentation, modeling methods for IHE content modules and profiling conventions, for frequently used standards, are maintained as [Appendix E](#) to the *IHE Technical Frameworks General Introduction*. Methods described include the standards conventions DICOM, HL7 v2.x, HL7 Clinical Document Architecture (CDA) Documents, etc. These conventions are critical to understanding this volume and should be reviewed prior to reading this text.

2.2 Additional Standards Profiling Conventions

This section defines profiling conventions for standards which are not described in the [IHE Technical Frameworks General Introduction](#).

No additional for Radiation Oncology.

3 Content Modules Overview and Terminology

In the future, an appendix to the *IHE Technical Frameworks General Introduction* will provide and an overview of Content Modules. In the interim, information may be available on the IHE wiki at <http://wiki.ihe.net/index.php?title=Profiles>

4 Reserved for domain specific content

Intentionally left blank

5 IHE Namespaces, Concept Domains, and Vocabularies

450 This section references the namespaces, concept domains, and identifiers defined or referenced by the IHE RO Technical Framework, and the vocabularies defined or referenced herein.

5.1 IHE Radiation Oncology Namespaces

No namespaces are defined.

5.2 IHE Radiation Oncology Concept Domains

No concept domains are defined.

455 5.3 IHE Radiation Oncology Format Codes and Vocabularies

The following vocabularies are referenced in the IHE RO Technical Framework. An extensive list of registered vocabularies can be found at <http://hl7.amg-hq.net/oid/frames.cfm>.

5.3.1 IHE Format Codes

460 For IHE Format Codes please see the IHE Format Codes wiki page at http://wiki.ihe.net/index.php/IHE_Format_Codes.

5.3.2 IHEActCode Vocabulary

- CCD ASTM/HL7 Continuity of Care Document
- CCR ASTM CCR Implementation Guide

465 The IHEActCode vocabulary is a small vocabulary of clinical acts that are not presently supported by the HL7 ActCode vocabulary. The root namespace (OID) for this vocabulary is 1.3.5.1.4.1.19376.1.5.3.2. These vocabulary terms are based on the vocabulary and concepts used in the CCR and CCD standards listed above.

Please see the IHEActCode Vocabulary at http://wiki.ihe.net/index.php/IHEActCode_Vocabulary.

470 5.3.3 IHERoleCode Vocabulary

The IHERoleCode vocabulary is a small vocabulary of role codes that are not presently supported by the HL7 Role Code vocabulary. The root namespace (OID) for this vocabulary is 1.3.5.1.4.1.19376.1.5.3.3.

475 Please see the IHERoleCode Vocabulary at http://wiki.ihe.net/index.php/IHERoleCode_Vocabulary

6 IHE RO HL7 V3 CDA Content Modules

No Content Modules defined.

7 IHE RO DICOM Content Modules

480 7.1 Conventions

The conventions of Volume 2, Section 2.2 DICOM Usage Conventions apply unless otherwise stated in the following.

7.1.1 Scope of Requirements

485 Requirements apply to all profiles which make use of the content definitions by referencing sections of this Volume. However where the uses cases covered by a profile need a different requirements, the profile may specify deviations from the definition here. This allows re-use of content definitions even in cases where only few adaptations are needed. It eliminates the need to duplicate the definitions, when the content requirements are shared in their majority and only a small number of deviations are indicated.

490 7.1.2 Requirements Definitions

Each content module has a list of attributes requirements. In any case, the requirements specified in the referenced DICOM Standard do apply.

495 Attributes not listed may or may not be present along the definition of the DICOM Standard. The producer may provide such attributes, but the receiver is not required to interpret them. Thus, it is not an error to include more than is asked for, but it is an error to reject a content based on additional attributes present.

Attributes, which may or may not be present by definition in the DICOM Standard, but shall not present under the definition of IHE-RO will be included in the specification with a requirement to be absent.

500 Attribute requirements are only in effect when the enclosing sequence item is present. For example, a type 1 attribute can be left out of content IF the enclosing sequence is not required and is not present.

505 IHE and IHE-RO have defined requirements related to the support for and use of attributes in DICOM storage transactions by both Service Class Users (SCUs) and Service Class Providers (SCPs):

IOD Table

M / C / U	As defined in DICOM PS 3.3
R	The Module is defined as Conditional (C) or User Option (U) in DICOM. The Requirement is an IHE extension of the DICOM requirements, and the module shall be present.
RC	The Module is defined as Conditional (C) or User Option (U) in DICOM. The Requirement is an IHE extension of the DICOM requirements, and the module shall be present when the specified conditions apply.

Module Table

O	The attribute or its value is optional, i.e., in DICOM it is Type 2 or 3.
O+*	The attribute is optional, but additional constraints have been added. Note: The specification approach does not force a Type 2 or Type 3 value to become a Type 1 by stating O+.
R+	The Requirement is an IHE extension of the DICOM requirements, and the attribute shall be present, i.e., is Type 1, whereas the DICOM requirement may be Type 2 or 3.
RC+	The Requirement is an IHE extension of the DICOM requirements, and the attribute shall be present when the condition is satisfied, i.e., is Type 1C, whereas the DICOM requirement may be Type 2 or 3. If the condition is not fulfilled, the DICOM definitions apply. Note, that this means that the attribute may be present / have a value also in case the condition does not apply.
D	The requirements of DICOM apply unchanged, but the attribute needs to be displayed.
-	No IHE extension of the DICOM requirements is defined. The attribute is listed for better readability or similar purpose.
X+	The attribute information is required to be absent. DICOM Type 2 attributes shall be present with no value. DICOM Type 3 attributes shall be absent.

510

7.1.3 Display Requirements

An asterisk (*) appearing on the attribute requirements indicates that the attribute does NOT need to be displayed

7.2 General Definitions515 **7.2.1 Character Sets****7.2.1.1 Support of Character Sets other than ISO-IR 100**

All actors shall support at least the Default Character Set and ISO-IR 100 (Latin-1) in all transactions. Other character sets as specified in Specific Character Set (0008,0005) shall be supported along the specification of the conformance statements of the involved actors.

520 Especially that means the following:

- It shall be possible for all actors involved in a transaction to use those character sets in their communication which all actors support along their conformance statements.
- When there are no character sets shared across all actors, ISO-RO 100 shall be used.

7.2.2 Propagation of Common Patient Information

525

Attribute (Tag)	CT Image	RT Structure Set	Geometric RT Plan	Dosimetric RT Plan	RT Dose	RT Treatment Record
Patient's Name (0010,0010)	Source	Copy	Copy	Copy	Copy	Copy
Patient ID (0010,0020)	Source	Copy	Copy	Copy	Copy	Copy
Patient's Birth Date (0010,0030)	Source	Copy	Copy	Copy	Copy	Copy
Patient's Sex (0010,0040)	Source	Copy	Copy	Copy	Copy	Copy

Interoperable exchange requires consistent patient information. Reasonable effort should be made to reconcile inconsistent patient information.

Inconsistent data received by downstream actors must be handled safely.

530 7.2.3 Study Handling

It is recommended that a new Study is created for the RT Structure Set.

The RT Structure Set may copy the Study IE of the treatment planning image series.

DICOM objects that are created based on this RT Structure Set instance or further derived instances should copy the Study IE of their predecessors.

535 If changes to Study-related attributes are required in the RT workflow, a new Study IE shall be created.

Note 1: If one copies the Study Instance UID, no study level attributes may be altered.

7.2.4 Frame of Reference Handling

540 All DICOM objects based on a planning image Set shall copy the Frame of Reference Module values from this image set.

7.3 IOD Definitions

This section defines each DICOM IOD used in the IHE Radiation Oncology domain in detail, specifying the standards used and the information defined.

7.3.1 Prescription IODs

545 *This section is present only to convey the envisioned section numbering.*

7.3.2 Plan IODs

7.3.2.1 Technique Specific RT Plan IODs

7.3.2.1.1 RT Plan IOD for Photon External Beam in Planning State

7.3.2.1.1.1 Referenced Standards

550 DICOM 2018e PS 3.3

7.3.2.1.1.2 IOD Definition

IE	Module	Reference	Usage	IHE-RO Usage
Patient	Patient	C.7.1.1	M	M See Section 7.4.1.1.1
	Clinical Trial Subject	C.7.1.3	U	U
Study	General Study	C.7.2.1	M	M See Section 7.4.1.2.1
	Patient Study	C.7.2.2	U	U
	Clinical Trial Study	C.7.2.3	U	U
Series	RT Series	C.8.8.1	M	M See Section 7.4.1.4.1
	Clinical Trial Series	C.7.3.2	U	U
Frame of Reference	Frame of Reference	C.7.4.1	U	R See Section 7.4.1.7.1
Equipment	General Equipment	C.7.5.1	M	M See Section 7.4.1.5.1
Plan	RT General Plan	C.8.8.9	M	M See Section 7.4.3.1.1
	RT Prescription	C.8.8.10	U	R See Section 7.4.3.2.1
	RT Tolerance Tables	C.8.8.11	U	U
	RT Patient Setup	C.8.8.12	U	R See Section 7.4.5.3.1
	RT Fraction Scheme	C.8.8.13	U	R See Section 7.4.3.3.1 and 7.4.3.3.2
	RT Beams	C.8.8.14	C - Required if RT Fraction Scheme Module exists and Number of Beams (300A,0080) is greater than zero for one or more fraction groups	R Definitions see below

IE	Module	Reference	Usage	IHE-RO Usage
	RT Brachy Application Setups	C.8.8.15	C - Required if RT Fraction Scheme Module exists and Number of Brachy Application Setups (300A,00A0) is greater than zero for one or more fraction groups	Absent
	Approval	C.8.8.16	U	R
	SOP Common	C.12.1	M	M See Section 7.4.1.6.1

RT Beams Module is defined as follows:

Beam Content Type	Section
Basic Static Beam	7.4.4.1.1
Basic Static MLC Beam	7.4.4.1.2
Arc Beam	7.4.4.1.3
MLC Fixed Aperture Arc Beam	7.4.4.1.4
MLC Variable Aperture Arc Beam	7.4.4.1.5
Hard Wedge Beam	7.4.4.1.6
Virtual Wedge Beam	7.4.4.1.7
Motorized Wedge Beam	7.4.4.1.8
Static Electron Beam	7.4.4.1.9
Step & Shoot Beam	7.4.4.1.10
Sliding Window Beam	7.4.4.1.11
IMAT/VMAT Beam	7.4.4.1.12
Photon Applicator Beam	7.4.4.1.13
Photon Applicator Arc Beam	7.4.4.1.14

555

7.3.2.1.2 RT Plan IOD for Photon External Beam in Delivery State

This section is present only to convey the envisioned section numbering.

7.3.2.2 RT Plan IOD for General Use

7.3.2.2.1 RT Plan IOD from Dosimetric Planning

560 **7.3.2.2.1.1 Referenced Standards**

DICOM 2018d Edition PS 3.3

7.3.2.2.1.2 IOD Definition

IE	Module	Reference	Usage	IHE-RO Usage
Patient	Patient	C.7.1.1	M	M See Section 7.4.1.1.1
	Clinical Trial Subject	C.7.1.3	U	U
Study	General Study	C.7.2.1	M	M See Section 7.4.1.2.1
	Patient Study	C.7.2.2	U	U
	Clinical Trial Study	C.7.2.3	U	U
Series	RT Series	C.8.8.1	M	M See Section 7.4.1.4.1
	Clinical Trial Series	C.7.3.2	U	U
Frame of Reference	Frame of Reference	C.7.4.1	U	M See Section 7.4.1.7.1
Equipment	General Equipment	C.7.5.1	M	M See Section 7.4.1.5.1
Plan	RT General Plan	C.8.8.9	M	M See Section 7.4.3.1.1
	RT Prescription	C.8.8.10	U	R See Section 7.4.3.2.1
	RT Tolerance Tables	C.8.8.11	U	U
	RT Patient Setup	C.8.8.12	U	R See below
	RT Fraction Scheme	C.8.8.13	U	R See Section 7.4.3.3.3
	RT Beams	C.8.8.14	C - Required if RT Fraction Scheme Module exists and Number of Beams (300A,0080) is greater than zero for one or more fraction groups	R Shall be present
	RT Brachy Application Setups	C.8.8.15	C - Required if RT Fraction Scheme Module exists and Number of Brachy Application Setups (300A,00A0) is greater than zero for one or more fraction groups	N/A
	Approval	C.8.8.16	U	M
	SOP Common	C.12.1	M	M
Common Instance Reference	C.12.2	U	C – Required if reference information is available	

RT Patient Setup Module is defined as follows:

Patient Setup Option	Section
Base Setup	See Section 7.4.3.4.1
Feet First Setup	See Section 7.4.3.4.2
Reoriented Setup	See Section 7.4.3.4.3
Decubitus Setup	See Section 7.4.3.4.4

565

7.3.2.2.2 RT Plan IOD for Dose Composition

This section is present only to convey the envisioned section numbering.

7.3.2.2.3 RT Plan IOD for Consistent Dose Tracking

This section is present only to convey the envisioned section numbering.

570 **7.3.2.2.4 RT Ion Plan IOD from Dosimetric Planning**

7.3.2.2.4.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.3.2.2.4.2IOD Definition

IE	Module	Reference	Usage	IHE-RO Usage
Patient	Patient	C.7.1.1	M	M See Section 7.4.1.1.1
	Clinical Trial Subject	C.7.1.3	U	U
Study	General Study	C.7.2.1	M	M See Section 7.4.1.2.1
	Patient Study	C.7.2.2	U	U
	Clinical Trial Study	C.7.2.3	U	U
Series	RT Series	C.8.8.1	M	M See Section 7.4.1.4.1
	Clinical Trial Series	C.7.3.2	U	U
Frame of Reference	Frame of Reference	C.7.4.1	U	M See Section 7.4.1.7.1
Equipment	General Equipment	C.7.5.1	M	M See Section 7.4.1.5.1
Plan	RT General Plan	C.8.8.9	M	M See Section 7.4.3.1.1
	RT Prescription	C.8.8.10	U	R See Section 7.4.3.2.1
	RT Ion Tolerance Tables	C.8.8.24	U	U
	RT Patient Setup	C.8.8.12	U	R See below
	RT Fraction Scheme	C.8.8.13	U	R See Section 7.4.3.3.3
	RT Ion Beams	C.8.8.25	C - Required if RT Fraction Scheme Module exists and Number of Beams (300A,0080) is greater than zero for one or more fraction groups	R Shall be present
	Approval	C.8.8.16	U	M
	SOP Common	C.12.1	M	M
	Common Instance Reference	C.12.2	U	C – Required if reference information is available

575 **RT Patient Setup Module is defined as follows:**

Patient Setup Option	Section
Base Setup	See Section 7.4.3.4.1

Patient Setup Option	Section
Feet First Setup	See Section 7.4.3.4.2
Reoriented Setup	See Section 7.4.3.4.3
Decubitus Setup	See Section 7.4.3.4.4

7.3.2.2.5 RT Plan IOD from Geometric Planning

7.3.2.2.5.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.3.2.2.5.2 Definition

IE	Module	Reference	Usage	IHE-RO Usage
Patient	Patient	C.7.1.1	M	M See Section 7.4.1.1.1
	Clinical Trial Subject	C.7.1.3	U	U
Study	General Study	C.7.2.1	M	M See Section 7.4.1.2.1
	Patient Study	C.7.2.2	U	U
	Clinical Trial Study	C.7.2.3	U	U
Series	RT Series	C.8.8.1	M	M See Section 7.4.1.4.1
	Clinical Trial Series	C.7.3.2	U	U
Frame of Reference	Frame of Reference	C.7.4.1	U	M See Section 7.4.1.7.1
Equipment	General Equipment	C.7.5.1	M	M See Section 7.4.1.5.1
Plan	RT General Plan	C.8.8.9	M	M See Section 7.4.3.1.1
	RT Prescription	C.8.8.10	U	U
	RT Tolerance Tables	C.8.8.11	U	U
	RT Patient Setup	C.8.8.12	U	R See below
	RT Fraction Scheme	C.8.8.13	U	U
	RT Beams	C.8.8.14	C - Required if RT Fraction Scheme Module exists and Number of Beams (300A,0080) is greater than zero for one or more fraction groups	R See Section 7.4.4.4.1 (Can be excluded for zero beams with non-isocentric model)
	RT Brachy Application Setups	C.8.8.15	C - Required if RT Fraction Scheme Module exists and Number of Brachy Application Setups (300A,00A0) is greater than zero for one or more fraction groups	N/A
	Approval	C.8.8.16	U	M
	SOP Common	C.12.1	M	M
Common Instance Reference	C.12.2	U	C – Required if reference information is available	

580 **7.3.3 Image IOD**

7.3.3.1 RT Image

This section is present only to convey the envisioned section numbering.

7.3.3.2 CT Image

7.3.3.2.1 CT Image in Planning State

585 *This section is present only to convey the envisioned section numbering.*

7.3.3.2.2 CT Image in Delivery State

This section is present only to convey the envisioned section numbering.

7.3.3.2.3 CT Image for General Use

7.3.3.2.3.1 Referenced Standards

590 DICOM 2018d Edition PS 3.3

7.3.3.2.3.2 IOD Definition

IE	Module	Reference	Usage	IHE-RO Usage
Patient	Patient	C.7.1.1	M	M
	Clinical Trial Subject	C.7.1.3	U	U
Study	General Study	C.7.2.1	M	M
	Patient Study	C.7.2.2	U	U
	Clinical Trial Study	C.7.2.3	U	U
Series	General Series	C.7.3.1	M	M See below
	Clinical Trial Series	C.7.3.2	U	U
Frame of Reference	Frame of Reference	C.7.4.1	M	M
Equipment	General Equipment	C.7.5.1	M	M
Image	General Image	C.7.6.1	M	R
	Image Plane	C.7.6.2	M	R See below
	Image Pixel	C.7.6.3	M	M

IE	Module	Reference	Usage	IHE-RO Usage
	Contrast/Bolus	C.7.6.4	C - Required if contrast media was used in this image	C - Required if contrast media was used in this image
	Device	C.7.6.12	U	U
	Specimen	C.7.6.22	U	U
	CT Image	C.8.2.1	M	M
	Overlay Plane	C.9.2	U	U
	VOI LUT	C.11.2	U	U
	SOP Common	C.12.1	M	M
	Common Instance Reference	C.12.2	U	C – Required if reference information is available

General Series Module is defined as follows:

Image Orientation Option	Section
Base Setup	See Section 7.4.1.3.1
Feet First Setup	See Section 7.4.1.3.2
Decubitus Setup	See Section 7.4.1.3.3

595

Image Plane Module is defined as follows:

Image Orientation Option	Section
Base Setup	See Section 7.4.6.2.1
Decubitus Setup	See Section 7.4.6.2.2

7.3.4 RT Structure Set IOD

7.3.4.1 RT Structure Set for General Use

600 **7.3.4.1.1 RT Structure Set for Basic Interoperability**

7.3.4.1.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.3.4.1.1.2IOD Definition

IE	Module	Reference	Usage	IHE-RO Usage
Patient	Patient	C.7.1.1	M	M See Section 7.4.1.1.1
	Clinical Trial Subject	C.7.1.3	U	U
Study	General Study	C.7.2.1	M	M See Section 7.4.1.2.1
	Patient Study	C.7.2.2	U	U
	Clinical Trial Study	C.7.2.3	U	U
Series	RT Series	C.8.8.1	M	M See Section 7.4.1.4.1
	Clinical Trial Series	C.7.3.2	U	U
Frame of Reference	Frame of Reference	C.7.4.1	U	M See Section 7.4.1.7.1
Equipment	General Equipment	C.7.5.1	M	M See Section 7.4.1.5.1
Structure Set	Structure Set	C.8.8.5	M	M See Section 7.4.8.3.1
	ROI Contour	C.8.8.6	M	R See below
	RT ROI Observation	C.8.8.8	M	M See Section 7.4.8.1.1
	Approval	C.8.8.16	U	U
	SOP Common	C.12.1	M	M
	Common Instance Reference	C.12.2	U	C – Required if reference information is available

605 **ROI Contour Module is defined as follows:**

Contouring Option	Section
On-slice contouring	See Section 7.4.8.2.1
Off-slice contouring	See Section 7.4.8.2.2

7.3.4.1.2 RT Structure Set Multi Modality Content

7.3.4.1.2.1 Referenced Standards

DICOM 2018d Edition PS 3.3

610

7.3.4.1.2.2 IOD Definition

IE	Module	Reference	Usage	IHE-RO Usage
Patient	Patient	C.7.1.1	M	M See Section 7.4.1.1.1
	Clinical Trial Subject	C.7.1.3	U	U
Study	General Study	C.7.2.1	M	M See Section 7.4.1.2.1
	Patient Study	C.7.2.2	U	U
	Clinical Trial Study	C.7.2.3	U	U
Series	RT Series	C.8.8.1	M	M See Section 7.4.1.4.1
	Clinical Trial Series	C.7.3.2	U	U
Frame of Reference	Frame of Reference	C.7.4.1	U	M See Section 7.4.1.7.1
Equipment	General Equipment	C.7.5.1	M	M See Section 7.4.1.5.1
Structure Set	Structure Set	C.8.8.5	M	M See Section 7.4.8.3.2
	ROI Contour	C.8.8.6	M	R See Section 7.3.4.1.1
	RT ROI Observation	C.8.8.8	M	M See Section 7.4.8.1.1
	Approval	C.8.8.16	U	U
	SOP Common	C.12.1	M	M
	Common Instance Reference	C.12.2	U	C – Required if reference information is available

7.3.5 RT Dose IOD

7.3.5.1 RT Dose IOD for General Use

7.3.5.1.1 RT Dose from Dosimetric Planning

615 **7.3.5.1.1.1 Referenced Standards**

DICOM 2018d Edition PS 3.3

7.3.5.1.1.2 IOD Definition

IE	Module	Reference	Usage	IHE-RO Usage
Patient	Patient	C.7.1.1	M	M See Section 7.4.1.1.1
	Clinical Trial Subject	C.7.1.3	U	U
Study	General Study	C.7.2.1	M	M See Section 7.4.1.2.1
	Patient Study	C.7.2.2	U	U
	Clinical Trial Study	C.7.2.3	U	U
Series	RT Series	C.8.8.1	M	M See Section 7.4.1.4.1
	Clinical Trial Series	C.7.3.2	U	U
Frame of Reference	Frame of Reference	C.7.4.1	M	M See Section 7.4.1.7.1
Equipment	General Equipment	C.7.5.1	M	M See Section 7.4.1.5.1
Dose	General Image	C.7.6.1	C - Required if dose data contains grid-based doses.	M
	Image Plane	C.7.6.2	C - Required if dose data contains grid-based doses.	R See Section 7.4.13.1.1
	Image Pixel	C.7.6.3	C - Required if dose data contains grid-based doses.	M
	Multi-Frame	C.7.6.6	C - Required if dose data contains grid-based doses and pixel data is multi-frame data.	R See Section 7.4.13.2.1
	Overlay Plane	C.9.2	U	U
	Multi-Frame Overlay	C.9.3	U	U
	Modality LUT	C.11.1	U	U
	RT Dose	C.8.8.3	M	M See Section 7.4.13.3.1

IE	Module	Reference	Usage	IHE-RO Usage
	RT DVH	C.8.8.4	U	RC Required for transactions [RO-BRTO-II-3] and [RO-BRTO-II-4] See Section 7.4.13.4.1
	Structure Set	C.8.8.5	C - Required if dose data contains dose points or isodose curves	Outside the scope of this profile.
	ROI Contour	C.8.8.6	C - Required if dose data contains dose points or isodose curves	Outside the scope of this profile.
	RT Dose ROI	C.8.8.7	C - Required if dose data contains dose points or isodose curves	Outside the scope of this profile.
	SOP Common	C.12.1	M	M
	Common Instance Reference	C.12.2	U	C – Required if reference information is available

7.3.6 Treatment Record

620 7.3.7 Reporting IOD

This section is present only to convey the envisioned section numbering.

7.3.8 ROI Dictionary IOD

This section is present only to convey the envisioned section numbering.

7.3.9 Workflow IOD

625 *This section is present only to convey the envisioned section numbering.*

7.3.10 Spatial Registration IOD

7.3.10.1 Spatial Registration IOD for General Use

7.3.10.1.1 Spatial Registration IOD Base Content

7.3.10.1.1.1 Referenced Standards

630 DICOM 2018d PS 3.3

7.3.10.1.1.2 IOD Definition

IE	Module	Reference	Usage	IHE-RO Usage
Patient	Patient	C.7.1.1	M	M See Section 7.4.1.1.1
	Clinical Trial Subject	C.7.1.3	U	U
Study	General Study	C.7.2.1	M	M
	Patient Study	C.7.2.2	U	U
	Clinical Trial Study	C.7.2.3	U	U
Series	General Series	C.7.3.1	M	M
	Clinical Trial Series	C.7.3.2	U	U
	Spatial Registration Series	C.20.1	M	M
Frame of Reference	Frame of Reference	C.7.4.1	M	M
Equipment	General Equipment	C.7.5.1	M	M
Spatial Registration	Spatial Registration	C.20.2	M	R See Section 7.4.10.1.1
	Common Instance Reference	C.12.2	M	M
	SOP Common	C.12.1	M	M

7.4 Module Definitions

635 This section defines each DICOM Module used in the IHE Radiation Oncology domain in detail, specifying the standards used and the information defined.

7.4.1 General Modules

7.4.1.1 Patient Module

7.4.1.1.1 Patient Module Base Content

7.4.1.1.1.1 Referenced Standards

640 DICOM 2018d Edition PS 3.3

7.4.1.1.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Patient's Name	(0010,0010)	R+	IHE requires that this element be present. This element is one of the primary patient identifying elements, and as such, all DICOM objects with the same Study Instance UID, must have the same value in this element. Equipment which creates new series based on other series (i.e., resampled series, new structure sets, plans, etc.) must preserve the value of this element to adhere to this profile.
Patient ID	(0010,0020)	R+	See Patient's Name (0010,0010) See also RAD TF-2: A.3
Patient's Birth Date	(0010,0030)	O+	See Patient's Name (0010,0010) See also RAD TF-2: A.3
Patient's Sex	(0010,0040)	O+	See Patient's Name (0010,0010) See also RAD TF-2: A.3

7.4.1.2 Study Module

645 7.4.1.2.1 General Study Module Base Content

7.4.1.2.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.1.2.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Study Instance UID	(0020,000D)	RC+*	IHE requires that this value be preserved in the following cases: If a set of images are resampled and re-exported. This new set of images will be a new series. This series will belong to the same study and will have the same study date. This is to facilitate grouping the images in a PACS. All other study level attributes mentioned in this table shall be preserved based on their existence, especially meaning to preserve an empty attribute value. Equipment which creates new series based on other series (i.e., resampled series, new structure sets, plans, etc.) must preserve the value of this element to adhere to this profile (see Section 7.2.3).
Study Date	(0008,0020)	RC+	[See Study Instance UID (0020,000D)]
Study Time	(0008,0030)	RC+	[See Study Instance UID (0020,000D)]
Study ID	(0020,0010)	RC+	[See Study Instance UID (0020,000D)]
Accession Number	(0008,0050)	RC+	[See Study Instance UID (0020,000D)]
Study Description	(0008,1030)	O+	[See Study Instance UID (0020,000D)]

650 **7.4.1.3 General Series Module**

7.4.1.3.1 General Series Module Base Content

7.4.1.3.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.1.3.1.2 Module Definition

655

Attribute	Tag	Type	Attribute Note
Series Date	(0008,0021)	RC+	Must be used and preserved, if present. If the producer creates a new series must be defined.
Series Time	(0008,0031)	RC+	Must be used and preserved, if present. If the producer creates a new series must be defined.
Patient Position	(0018,5100)	R+	Shall be one of {HFS, HFP}.

7.4.1.3.2 General Series Module Feet First

7.4.1.3.2.1 Referenced Standards

DICOM 2018d Edition PS 3.3

660 **7.4.1.3.2.2 Module Definition**

Attribute	Tag	Type	Attribute Note
Series Date	(0008,0021)	RC+	Must be used and preserved, if present. If the producer creates a new series must be defined.
Series Time	(0008,0031)	RC+	Must be used and preserved, if present. If the producer creates a new series must be defined.
Patient Position	(0018,5100)	R+	Shall be one of {HFS, FFS, HFP, FFP}.

7.4.1.3.3 General Series Module Decubitus

7.4.1.3.3.1 Referenced Standards

DICOM 2018d Edition PS 3.3

665

7.4.1.3.3.2 Module Definition

Attribute	Tag	Type	Attribute Note
Series Date	(0008,0021)	RC+	Must be used and preserved, if present. If the producer creates a new series must be defined.
Series Time	(0008,0031)	RC+	Must be used and preserved, if present. If the producer creates a new series must be defined.
Patient Position	(0018,5100)	R+	Shall be one of {HFS, FFS, HFP, FFP, HFDL, HFDR, FFDL, FFDL}.

7.4.1.4 RT Series Module

670 7.4.1.4.1 RT Series Module Base Content

7.4.1.4.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.1.4.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Series Date	(0008,0021)	RC+	Must be used and preserved, if present. If the producer creates a new series must be defined.
Series Time	(0008,0031)	RC+	Must be used and preserved, if present. If the producer creates a new series must be defined.

675

7.4.1.5 Equipment Module

7.4.1.5.1 General Equipment Module Base Content

7.4.1.5.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

680 7.4.1.5.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Manufacturer	(0008,0070)	R+*	IHE requires that this element be present, and should contain the manufacturer of the equipment creating the image, structure set, plan, or dose.

Attribute	Tag	Type	Attribute Note
			If the equipment is storing and forwarding information, the value of this element shall be preserved. If a new plan is created from a previous plan, the manufacturer of the equipment producing the new plan shall insert their identifier in this element. If a new structure set is created from a previous structure set, the manufacturer of the equipment producing the new structure set shall insert their identifier in this element.
Manufacturer's Model Name	(0008,1090)	R+*	If an application resamples and re-exports a series of CT images, or modifies an instance then this element must be present, and must contain the model name of the equipment doing the resampling.
Software Versions	(0018,1020)	R+*	Must be present.

7.4.1.6 SOP Common Module

7.4.1.6.1 SOP Common Module Base Content

685 7.4.1.6.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.1.6.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Specific Character Set	(0008,0005)	O+*	See Section 7.2.1
Instance Creation Date	(0008,0012)	R+	Shall be present.
Instance Creation Time	(0008,0013)	R+	Shall be present.

7.4.1.7 Frame of Reference Module

690 7.4.1.7.1 Frame of Reference Module Base Content

7.4.1.7.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

695

7.4.1.7.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Position Reference Indicator	(0020,1040)	O*	Equipment which creates new series based on other series (i.e., resampled series, new structure sets, plans, etc.) must preserve the value of this element to adhere to this profile.

7.4.1.8 General Image Module

7.4.1.8.1 General Image Module Base Content

700 7.4.1.8.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.1.8.1.2 Module Definition

See Treatment Delivery- Image Content (TDIC) Profile, Section 7.4.1.8.1.

7.4.2 Workflow-Related Modules

705 *This section is present only to convey the envisioned section numbering.*

7.4.3 General Plan-Related Modules

7.4.3.1 General Plan Module

7.4.3.1.1 General Plan Module Base Content

7.4.3.1.1.1 Referenced Standards

710 DICOM 2018d Edition PS 3.3

7.4.3.1.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
RT Plan Label	(300A,0002)	R+	The label which serves as the identification of the plan for the user.
RT Plan Date	(300A,0006)	R+	The date when the plan was last modified.
RT Plan Time	(300A,0007)	R+	The time when the plan was last modified.
RT Plan Geometry	(300A,000C)	R+*	Shall be PATIENT. This implies that the RT Structure Set exists and is referenced in the General Plan module.

7.4.3.2 RT Prescription Module

7.4.3.2.1 RT Prescription Module Base Content

715 **7.4.3.2.1.1 Referenced Standards**

DICOM 2018e Edition PS 3.3

7.4.3.2.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Dose Reference Sequence	(300A,0010)	R+*	See Dose reference requirements in the RT Beams module for the TPPC transactions.
> Dose Reference UID	(300A,0013)	R+*	
> Dose Reference Description	(300A,0016)	R+	
> Target Prescription Dose	(300A,0026)	O+	If present, shall be of the same type of dose as the Beam Dose Type (300A,0090) in the RT Fraction Scheme Module (see 7.4.3.3.2)

7.4.3.3 RT Fraction Scheme Module

720 **7.4.3.3.1 RT Fraction Scheme Module for Consistent Dose**

<i>This section is present only to convey the envisioned section numbering.</i>

7.4.3.3.2 RT Fraction Scheme Module for Delivery

7.4.3.3.2.1 Referenced Standards

DICOM 2018e Edition PS 3.3

725 **7.4.3.3.2.2 Module Definition**

Attribute	Tag	Type	Attribute Note
Fraction Group Sequence	(300A,0070)	R+*	Shall have only a single item in the sequence
> Number of Fractions Planned	(300A,0078)	R+	
> Referenced Beam Sequence	(300C,0004)	R+*	
>> Referenced Dose Reference UID	(300A,0083)	R+*	Identifies the Dose Reference specified by Dose Reference UID (300A,0013) in the Dose Reference Sequence (300A,0010) in the RT

Attribute	Tag	Type	Attribute Note
			Prescription Module which specifies the primary target for the current Beam. If present shall have a value that is present in the Dose Reference Sequence.
>> Beam Dose	(300A,0084)	-/R+ R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it.
>> Beam Dose Specification Point	(300A,0082)	R+	
>> Beam Meterset	(300A,0086)	R+	
>> Beam Dose Type	(300A,0090)	R+	Shall be present

7.4.3.3.3 RT Fraction Scheme Module for Brachy

This section is present only to convey the envisioned section numbering.

730 7.4.3.3.4 RT Fraction Scheme Module Base Content

7.4.3.3.4.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.3.3.4.2 Module Definition

Attribute	Tag	Type	Attribute Note
Fraction Group Sequence	(300A,0070)	R+*	Shall have only a single item in the sequence
>Number of Brachy Application Setups	(300A,00A0)	R+*	Shall be 0. Brachytherapy is not supported in the BRTO Profile.

7.4.3.4 RT Patient Setup Module

735 7.4.3.4.1 RT Patient Setup Module Base Content

7.4.3.4.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.3.4.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Patient Setup Sequence	(300A,0180)	R+*	An actor must not rely on the presence of: Fixation Device Sequence Shielding Device Sequence Setup Device Sequence within the Patient Setup Sequence for proper operation.
>Patient Position	(0018,5100)	R+	Shall be one of {HFS, HFP}. In case of multiple Patient Setup items, it shall be the same.
>Setup Technique	(300A,01B0)	R+*	
>Table Top Vertical Setup Displacement	(300A,01D2)	O+*	If present, shall be consistent with Isocenter position. See note below.
>Table Top Longitudinal Setup Displacement	(300A,01D4)	O+*	If present, shall be consistent with Isocenter position. See note below.
>Table Top Longitudinal Setup Displacement	(300A,01D6)	O+*	If present, shall be consistent with Isocenter position. See note below.

Note: All items in the Patient Setup Sequence (300A,0180) shall use the same initial Setup Position.

740 **7.4.3.4.2 RT Patient Setup Module Feet First****7.4.3.4.2.1 Referenced Standards**

DICOM 2018d Edition PS 3.3

7.4.3.4.2.2 Module Definition

Attribute	Tag	Type	Attribute Note
Patient Setup Sequence	(300A,0180)	R+*	An actor must not rely on the presence of: Fixation Device Sequence Shielding Device Sequence Setup Device Sequence within the Patient Setup Sequence for proper operation.
>Patient Position	(0018,5100)	R+	Shall be one of {HFS, FFS, HFP, FFP}.
>Setup Technique	(300A,01B0)	R+*	
>Table Top Vertical Setup Displacement	(300A,01D2)	O+*	If present, shall be consistent with Isocenter position. See note below.
>Table Top Longitudinal Setup Displacement	(300A,01D4)	O+*	If present, shall be consistent with Isocenter position. See note below.
>Table Top Longitudinal Setup Displacement	(300A,01D6)	O+*	If present, shall be consistent with Isocenter position. See note below.

Note: All items in the Patient Setup Sequence (300A,0180) shall use the same initial Setup Position.

745 **7.4.3.4.3 RT Patient Setup Module Reoriented****7.4.3.4.3.1 Referenced Standards**

DICOM 2018d Edition PS 3.3

7.4.3.4.3.2 Module Definition

Attribute	Tag	Type	Attribute Note
Patient Setup Sequence	(300A,0180)	R+*	An actor must not rely on the presence of: Fixation Device Sequence Shielding Device Sequence Setup Device Sequence within the Patient Setup Sequence for proper operation.
>Patient Position	(0018,5100)	R+	The reoriented Patient Position for treatment shall correspond to the following pairs with respect of the Patient Position during image acquisition: HFS ↔ FFS or HFP ↔ FFP
>Setup Technique	(300A,01B0)	R+*	
>Table Top Vertical Setup Displacement	(300A,01D2)	O+*	If present, shall be consistent with Isocenter position. See note below.
>Table Top Longitudinal Setup Displacement	(300A,01D4)	O+*	If present, shall be consistent with Isocenter position. See note below.
>Table Top Longitudinal Setup Displacement	(300A,01D6)	O+*	If present, shall be consistent with Isocenter position. See note below.

Note: All items in the Patient Setup Sequence (300A,0180) shall use the same initial Setup Position.

750 **7.4.3.4.4 RT Patient Setup Module Decubitus****7.4.3.4.4.1 Referenced Standards**

DICOM 2018d Edition PS 3.3

7.4.3.4.4.2 Module Definition

Attribute	Tag	Type	Attribute Note
Patient Setup Sequence	(300A,0180)	R+*	An actor must not rely on the presence of: Fixation Device Sequence Shielding Device Sequence Setup Device Sequence within the Patient Setup Sequence for proper operation.
>Patient Position	(0018,5100)	R+	Shall be one of {HFS, FFS, HFP, FFP, HF DL, HF DR, FF DL, FF DR}.

Attribute	Tag	Type	Attribute Note
>Setup Technique	(300A,01B0)	R+*	
>Table Top Vertical Setup Displacement	(300A,01D2)	O+*	If present, shall be consistent with Isocenter position. See note below.
>Table Top Longitudinal Setup Displacement	(300A,01D4)	O+*	If present, shall be consistent with Isocenter position. See note below.
>Table Top Longitudinal Setup Displacement	(300A,01D6)	O+*	If present, shall be consistent with Isocenter position. See note below.

Note: All items in the Patient Setup Sequence (300A,0180) shall use the same initial Setup Position.

755 **7.4.4 Plan-Related Modules in Planning**

7.4.4.1 Specific RT Beam Type Specifications

7.4.4.1.1 RT Beams Module for Basic Static Beam

7.4.4.1.1.1 Referenced Standards

DICOM 2018e Edition PS 3.3

760 **7.4.4.1.1.2 Module Definition**

Attribute	Tag	Beam Technique	
		Basic Static	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be >= 1.
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be STATIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	

Attribute	Tag	Beam Technique	
		Basic Static	
		Presence	Specific Rules
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall be 2 jaws, MLC shall not be present
>> Leaf Position Boundaries	(300A,00BE)	O+*	NA (no MLC) May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be >= 1.
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0 or 1. If 1, see Compensator Beam Modifier.
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0, see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	O+	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall not be present
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	Shall be constant.
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be NONE.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.

Attribute	Tag	Beam Technique	
		Basic Static	
		Presence	Specific Rules
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

7.4.4.1.2 RT Beams Module for Basic Static MLC Beam

7.4.4.1.2.1 Referenced Standards

765 DICOM 2018e Edition PS 3.3

7.4.4.1.2.2 Module Definition

Attribute	Tag	Beam Technique	
		Basic Static MLC	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be >= 1.
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be STATIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall have at least 1 MLC

Attribute	Tag	Beam Technique	
		Basic Static MLC	
		Presence	Specific Rules
>> Leaf Position Boundaries	(300A,00BE)	R+*	Shall be present for MLCs May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be >= 1.
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0 or 1. If 1, see Compensator Beam Modifier.
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0, see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall not be present
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	Shall be constant.
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be NONE.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.

Attribute	Tag	Beam Technique	
		Basic Static MLC	
		Presence	Specific Rules
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

7.4.4.1.3 RT Beams Module for Arc Beam

770 7.4.4.1.3.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.1.3.2 Module Definition

Attribute	Tag	Beam Technique	
		Arc	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be ≥ 1 .
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be DYNAMIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall be 2 jaws, MLC shall not be present
>> Leaf Position Boundaries	(300A,00BE)	O+*	NA (no MLC) May or may not be present for jaws, may be ignored for jaws

Attribute	Tag	Beam Technique	
		Arc	
		Presence	Specific Rules
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be >= 1.
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0.
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0, see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2. Skip arcs are not tested in this transaction.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>>> Referenced Dose Reference Sequence	(300C,0050)	-/R+*	A TMS Actor is required to consume and process this value.
		R+/O+*	A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it Shall have at least one item for target dose accumulation.
>>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>>> Wedge Position Sequence	(300A,0116)	R+*	Shall not be present
>>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>> Leaf/Jaw Positions	(300A,011C)	R+*	
>>> Gantry Angle	(300A,011E)	R+*	
>>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be CW or CC for Control Point 0. Can be NONE for Control Point 1.
>>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.

Attribute	Tag	Beam Technique	
		Arc	
		Presence	Specific Rules
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

775 7.4.4.1.4 RT Beams Module for MLC Fixed Aperture Arc Beam

7.4.4.1.4.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.1.4.2 Module Definition

Attribute	Tag	Beam Technique	
		MLC Fixed Aperture Arc	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be ≥ 1 .
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be DYNAMIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall have at least 1 MLC.
>> Leaf Position Boundaries	(300A,00BE)	R+*	Shall be present for MLCs. May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be ≥ 1 .
> Treatment Delivery Type	(300A,00CE)	R+*	

Attribute	Tag	Beam Technique	
		MLC Fixed Aperture Arc	
		Presence	Specific Rules
> Number of Wedges	(300A,00D0)	R+*	Shall be 0.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0.
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2. Skip arcs are not tested in this transaction.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>>> Referenced Dose Reference Sequence	(300C,0050)	-/R+*	A TMS Actor is required to consume and process this value.
		R+/O+*	A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it Shall have at least one item for target dose accumulation.
>>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall not be present
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>> Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be CW or CC for Control Point 0. Can be NONE for Control Point 1.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

780 **7.4.4.1.5 RT Beams Module for MLC Variable Aperture Arc Beam****7.4.4.1.5.1 Referenced Standards**

DICOM 2018e Edition PS 3.3

7.4.4.1.5.2 Module Definition

Attribute	Tag	Beam Technique	
		MLC Variable Aperture Arc	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be ≥ 1 .
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be DYNAMIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall be 2 jaws, or at least 1 jaw and 1 MLC.
>> Leaf Position Boundaries	(300A,00BE)	R+*	Shall be present for MLCs. May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be ≥ 1 .
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0.
> Number of Boli	(300A, 00ED)	R+*	Shall be ≥ 0 . If > 0 , see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0 , see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	

Attribute	Tag	Beam Technique	
		MLC Variable Aperture Arc	
		Presence	Specific Rules
> Number of Control Points	(300A,0110)	R+*	If the Consumer has a limit, it shall document this and safely handle input that exceeds that limit.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>>> Referenced Dose Reference Sequence	(300C,0050)	-/R+*	A TMS Actor is required to consume and process this value.
		R+/O+*	A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it Shall have at least one item for target dose accumulation.
>>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>>> Wedge Position Sequence	(300A,0116)	R+*	Shall not be present
>>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>> Leaf/Jaw Positions	(300A,011C)	R+*	
>>>> Gantry Angle	(300A,011E)	R+*	
>>>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be constant (CW or CC) for all CP except last one. Can be NONE for final CP
>>>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>>>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>>>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>>>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>>>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

785

7.4.4.1.6 RT Beams Module for Hard Wedge Beam

7.4.4.1.6.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.1.6.2 Module Definition

790

Attribute	Tag	Beam Technique	
		Hard Wedge	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be ≥ 1 .
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be STATIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall have at least 2 jaws or at least 1 jaw and 1 MLC.
>>> Leaf Position Boundaries	(300A,00BE)	R+*	Shall be present for MLCs,. May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be ≥ 1 .
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 1.
>> Wedge Sequence	(300A,00D1)	R+*	Shall be present.
>> Wedge Type	(300A,00D3)	R+*	Shall be STANDARD (static)
>> Wedge ID	(300A,00D4)	R+	
>> Wedge Angle	(300A,00D5)	R+	
>> Wedge Orientation	(300A,00D8)	R+	
>> Source to Wedge Tray Distance	(300A,00DA)	R+	
> Number of Compensators	(300A,00E0)	R+*	Shall be 0 or 1. If 1, see Compensator Beam Modifier.
> Number of Boli	(300A, 00ED)	R+*	Shall be ≥ 0 . If > 0 , see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0 , see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.

Attribute	Tag	Beam Technique	
		Hard Wedge	
		Presence	Specific Rules
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+*	A TMS Actor is required to consume and process this value.
		R+/O+*	A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall be present and consistent with the Wedge Sequence (300A,00D1).
>>> Wedge Position	(300A,0118)	R+*	Shall be IN.
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	Shall be constant.
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be NONE.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List >(See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

7.4.4.1.7 RT Beams Module for Virtual Wedge Beam**7.4.4.1.7.1 Referenced Standards**

DICOM 2018e Edition PS 3.3

795

7.4.4.1.7.2 Module Definition

Attribute	Tag	Beam Technique	
		Virtual Wedge	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be ≥ 1 .
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be STATIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall have at least 2 jaws or at least 1 jaw and 1 MLC.
>>> Leaf Position Boundaries	(300A,00BE)	R+*	Shall be present for MLCs,. May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be ≥ 1 .
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 1 or 2. If 2, see Hard Wedge Beam Modifier.
>> Wedge Sequence	(300A,00D1)	R+*	Shall be present if number of wedges is non-zero
>>> Wedge Type	(300A,00D3)	R+*	Shall be DYNAMIC. Optional Hard Wedge shall be STANDARD
>>> Wedge ID	(300A,00D4)	R+	
>>> Wedge Angle	(300A,00D5)	RC+	Shall be present if Wedge Type (300A,00D3) is STANDARD. May be present otherwise.

Attribute	Tag	Beam Technique	
		Virtual Wedge	
		Presence	Specific Rules
>>Effective Wedge Angle	(300A,00DE)	RC+/O+*	Shall be present if Wedge Type (300A,00D3) is DYNAMIC. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it
>> Wedge Orientation	(300A,00D8)	R+	
>> Source to Wedge Tray Distance	(300A,00DA)	RC+	Shall be present if Wedge Type (300A,00D3) is STANDARD.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0 or 1. If 1, see Compensator Beam Modifier.
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0, see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall be present and consistent with the Wedge Sequence (300A,00D1).
>>> Wedge Position	(300A,0118)	R+*	Shall be IN.
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	Shall be constant.
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be NONE.

Attribute	Tag	Beam Technique	
		Virtual Wedge	
		Presence	Specific Rules
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

7.4.4.1.8 RT Beams Module for Motorized Wedge Beam

7.4.4.1.8.1 Referenced Standards

800 DICOM 2018e Edition PS 3.3

7.4.4.1.8.2 Module Definition

Attribute	Tag	Beam Technique	
		Motorized Wedge	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be ≥ 1 .
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be STATIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	

Attribute	Tag	Beam Technique	
		Motorized Wedge	
		Presence	Specific Rules
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall have at least 2 jaws or at least 1 jaw and 1 MLC.
>> Leaf Position Boundaries	(300A,00BE)	R+*	Shall be present for MLCs,. May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be >= 1.
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 1 or 2. If 2, see also Hard Wedge Beam Modifier
> Wedge Sequence	(300A,00D1)	R+*	Shall be present.
>> Wedge Type	(300A,00D3)	R+*	Shall be MOTORIZED. Optional Hard Wedge shall be STANDARD
>> Wedge ID	(300A,00D4)	R+	
>> Wedge Angle	(300A,00D5)	RC+	Shall be present if Wedge Type (300A,00D3) is STANDARD. May be present otherwise.
>> Wedge Orientation	(300A,00D8)	R+	
>> Source to Wedge Tray Distance	(300A,00DA)	RC+	Shall be present if Wedge Type (300A,00D3) is STANDARD.
>>Effective Wedge Angle	(300A,00DE)	RC+/O+*	Shall be present if Wedge Type (300A,00D3) is MOTORIZED. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it
> Number of Compensators	(300A,00E0)	R+*	Shall be 0 or 1. If 1, see Compensator Beam Modifier.
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0, see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 4.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	

Attribute	Tag	Beam Technique	
		Motorized Wedge	
		Presence	Specific Rules
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall be present and consistent with the Wedge Sequence (300A,00D1).
>>> Wedge Position	(300A,0118)	R+*	For Motorized Wedge, shall be IN for CPs 0 and 1, OUT for CPs 2 and 3. Shall be IN for optional Hard Wedge
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>> Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	Shall be constant.
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be NONE.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

7.4.4.1.9 RT Beams Module for Basic Static Electron Beam

805 7.4.4.1.9.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.1.9.2 Module Definition

Attribute	Tag	Beam Technique	
		Basic Static Electron	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be >= 1.
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be STATIC.
> Radiation Type	(300A,00C6)	R+*	Shall be ELECTRON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall be 2 jaws, MLC shall not be present
>>> Leaf Position Boundaries	(300A,00BE)	O+*	NA (no MLC) May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be >= 1.
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0 or 1. If 1, see Compensator Beam Modifier.
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0, see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	R+*	Shall contain 1 item.
>> Applicator ID	(300A,0108)	R+	
>> Applicator Type	(300A,0109)	R+*	
>>> Applicator Geometry Sequence	(300A,0431)	R+*	
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2.

Attribute	Tag	Beam Technique	
		Basic Static Electron	
		Presence	Specific Rules
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall not be present
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	Shall be constant.
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be NONE.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.
>> Source to Surface Distance	(300A,0130)	-/R+ R+/O+*	A TMS Actor is required to consume and process this value. A beam consumer/producer Actor (e.g., a TPS) may consume this value and is required to produce it if Patient Setup Technique (300A, 01B0) is FIXED_SSD.

Attribute	Tag	Beam Technique	
		Basic Static Electron	
		Presence	Specific Rules
>>Source to External Contour Distance	(300A,0132)	-/R+ R+/O+*	A TMS Actor is required to consume and process this value. A beam consumer/producer Actor (e.g., a TPS) may consume this value and is required to produce it if Patient Setup Technique (300A, 01B0) is FIXED_SSD.

810 7.4.4.1.10 RT Beams Module for Step & Shoot Beam

7.4.4.1.10.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.1.10.2 Module Definition

Attribute	Tag	Beam Technique	
		Step & Shoot	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be >= 1.
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be STATIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	At least 1 MLC shall be present

Attribute	Tag	Beam Technique	
		Step & Shoot	
		Presence	Specific Rules
>> Leaf Position Boundaries	(300A,00BE)	R+*	Shall be present for MLCs. May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be >= 1.
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0 or 1. If 1, see Hard Wedge Beam Modifier
> Number of Compensators	(300A,00E0)	R+*	Shall be 0..
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0, see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2n, where n is the number of unique field shapes composing the beam If the Consumer has a limit, it must document this and safely handle input that exceeds the limit
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	CP[0]=0.0 :: CP[2n + 1]=Cumulative Meterset Weight after completion of delivery of the field shape :: CP[2n+1] = CP[2n + 2]
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	See Hard Wedge Beam Modifier If present, may not be ignored
>>> Wedge Position	(300A,0118)	R+*	Shall be IN.

Attribute	Tag	Beam Technique	
		Step & Shoot	
		Presence	Specific Rules
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	Shall be constant.
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be NONE.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

815

7.4.4.1.11 RT Beams Module for Sliding Window Beam

7.4.4.1.11.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.1.11.2 Module Definition

820

Attribute	Tag	Beam Technique	
		Sliding Window	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be >= 1.
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be DYNAMIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	

Attribute	Tag	Beam Technique	
		Sliding Window	
		Presence	Specific Rules
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	At least 1 MLC shall be present
>> Leaf Position Boundaries	(300A,00BE)	R+*	Shall be present for MLCs. May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be ≥ 1 .
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0 or 1. If 1, see Hard Wedge Beam Modifier
> Number of Compensators	(300A,00E0)	R+*	Shall be 0..
> Number of Boli	(300A, 00ED)	R+*	Shall be ≥ 0 . If > 0 , see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0-8. If > 0 , see Block Beam Modifier.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall > 2 . If the Consumer has a limit, it must document this and safely handle input that exceeds the limit
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.

Attribute	Tag	Beam Technique	
		Sliding Window	
		Presence	Specific Rules
>> Wedge Position Sequence	(300A,0116)	R+*	See Hard Wedge Beam Modifier If present, may not be ignored
>>> Wedge Position	(300A,0118)	R+*	Shall be IN.
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	Shall be constant.
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be NONE.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

7.4.4.1.12 RT Beams Module for IMAT/VMAT Beam

7.4.4.1.12.1 Referenced Standards

DICOM 2018e Edition PS 3.3

825 7.4.4.1.12.2 Module Definition

Attribute	Tag	Beam Technique	
		IMAT/VMAT	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be >= 1.
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be DYNAMIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	

Attribute	Tag	Beam Technique	
		IMAT/VMAT	
		Presence	Specific Rules
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	At least 1 MLC shall be present
>> Leaf Position Boundaries	(300A,00BE)	R+*	Shall be present for MLCs. May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be >= 1.
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0..
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0.
> Applicator Sequence	(300A,0107)	R+*	Shall not be present.
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be > 2. If the Consumer has a limit, it must document this and safely handle input that exceeds the limit
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+* R+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be present as Nominal Dose Rate.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall not be present.

Attribute	Tag	Beam Technique	
		IMAT/VMAT	
		Presence	Specific Rules
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be constant (CW or CC) for all CP except the last CP, which can be NONE.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	If present, shall not be ignored.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

7.4.4.1.13 RT Beams Module for Photon Applicator Beam

7.4.4.1.13.1 Referenced Standards

830 DICOM 2018e Edition PS 3.3

7.4.4.1.13.2 Module Definition

Attribute	Tag	Beam Technique	
		Photon Applicator	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be ≥ 1 .
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be STATIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	
>> Fluence Mode	(3002,0051)	D	

Attribute	Tag	Beam Technique	
		Photon Applicator	
		Presence	Specific Rules
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall be 2 jaws, MLC shall not be present
>> Leaf Position Boundaries	(300A,00BE)	O+*	NA (no MLC) May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be >= 1.
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0.
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0.
> Applicator Sequence	(300A,0107)	R+*	Shall contain 1 item.
>> Applicator ID	(300A,0108)	R+	
>> Applicator Type	(300A,0109)	R+*	Shall be PHOTON_CIRC
>> Applicator Geometry Sequence	(300A,0431)	R+*	
>>> Applicator Aperture Shape	(300A,0432)	R+	Shall be SYM_CIRCULAR
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.

Attribute	Tag	Beam Technique	
		Photon Applicator	
		Presence	Specific Rules
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall not be present
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	Shall be constant.
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be NONE.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

7.4.4.1.14 RT Beams Module for Photon Applicator Arc Beam

835 7.4.4.1.14.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.1.14.2 Module Definition

Attribute	Tag	Beam Technique	
		Photon Applicator Arc	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Beam Number	(300A,00C0)	R+*	Shall be ≥ 1 .
> Beam Name	(300A,00C2)	R+	
> Beam Type	(300A,00C4)	R+*	Shall be DYNAMIC.
> Radiation Type	(300A,00C6)	R+*	Shall be PHOTON.
> High-Dose Technique Type	(300A,00C7)	O+*	If present, must be handled safely
> Primary Fluence Mode Sequence	(3002,0050)	R+*	

Attribute	Tag	Beam Technique	
		Photon Applicator Arc	
		Presence	Specific Rules
>> Fluence Mode	(3002,0051)	D	
>> Fluence Mode ID	(3002,0052)	D	
> Treatment Machine Name	(300A,00B2)	R+*	Shall be constant.
> Primary Dosimeter Unit	(300A,00B3)	R+	Shall be MU.
> Source-Axis Distance	(300A,00B4)	R+*	
> Beam Limiting Device Sequence	(300A,00B6)	R+*	
>> RT Beam Limiting Device Type	(300A,00B8)	R+*	Shall be 2 jaws, MLC shall not be present
>> Leaf Position Boundaries	(300A,00BE)	O+*	NA (no MLC) May or may not be present for jaws, may be ignored for jaws
> Referenced Patient Setup Number	(300C,006A)	R+*	Shall be >= 1.
> Treatment Delivery Type	(300A,00CE)	R+*	
> Number of Wedges	(300A,00D0)	R+*	Shall be 0.
> Number of Compensators	(300A,00E0)	R+*	Shall be 0.
> Number of Boli	(300A, 00ED)	R+*	Shall be >= 0. If > 0, see Bolus Beam Modifier.
> Number of Blocks	(300A,00F0)	R+*	Shall be 0.
> Applicator Sequence	(300A,0107)	R+*	Shall contain 1 item.
>> Applicator ID	(300A,0108)	R+	
>> Applicator Type	(300A,0109)	R+*	Shall be PHOTON_CIRC.
>> Applicator Geometry Sequence	(300A,0431)	R+*	
>>> Applicator Aperture Shape	(300A,0432)	R+	Shall be SYM_CIRCULAR
> Final Cumulative Meterset Weight	(300A,010E)	R+*	
> Number of Control Points	(300A,0110)	R+*	Shall be 2.
> Control Point Sequence	(300A,0111)	R+*	
>> Cumulative Meterset Weight	(300A,0134)	R+	
>> Referenced Dose Reference Sequence	(300C,0050)	-/R+* R+/O+*	A TMS Actor is required to consume and process this value. A beam producer/consumer Actor (e.g., a TPS) may consume this value and is required to produce it. Shall have at least one item for target dose accumulation.
>>> Cumulative Dose Reference Coefficient	(300A,010C)	R+*	Shall be present.

Attribute	Tag	Beam Technique	
		Photon Applicator Arc	
		Presence	Specific Rules
>> Nominal Beam Energy	(300A,0114)	R+	Shall be constant.
>> Dose Rate Set	(300A,0115)	R+	Shall be constant.
>> Wedge Position Sequence	(300A,0116)	R+*	Shall not be present
>> Beam Limiting Device Position Sequence	(300A,011A)	R+*	Shall be consistent with the Beam Limiting Device Sequence (300A,00B6).
>>>Leaf/Jaw Positions	(300A,011C)	R+*	
>> Gantry Angle	(300A,011E)	R+*	
>> Gantry Rotation Direction	(300A,011F)	R+*	Shall be CW or CC for Control Point 0 Can be NONE for Control Point 1.
>> Gantry Pitch Angle	(300A,014A)	O+*	If not present, shall be assumed to be in the zero position. If present, shall be zero.
>> Gantry Pitch Rotation Direction	(300A,014C)	O+*	If present, shall be NONE.
>> Beam Limiting Device Angle	(300A,0120)	R+*	Shall be constant.
>> Beam Limiting Device Rotation Direction	(300A,0121)	R+*	Shall be NONE.
< Insert Control Point Sequence Fixed Attributes List > (See Section 7.4.4.2.1)			
>> Isocenter Position	(300A,012C)	R+	Shall be constant for all CPs.

840 **7.4.4.2 General Beam Attribute Specifications**

7.4.4.2.1 Control Point Fixed Attribute List Base Content

7.4.4.2.1.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.2.1.2 Required Attributes

845 The list of attributes and requirements below shall be included in all TPPC transactions as noted in the RT Beam Module specification associated with those transactions.

Attribute	Tag	Control Point Sequence	
		Fixed Attributes	
		Presence	Specific Rules
>> Patient Support Angle	(300A,0122)	R+*	Shall be constant.
>> Patient Support Rotation Direction	(300A,0123)	R+*	Shall be NONE.

Attribute	Tag	Control Point Sequence	
		Fixed Attributes	
		Presence	Specific Rules
>> Table Top Eccentric Axis Distance	(300A,0124)	O+*	If present, shall be constant.
>> Table Top Eccentric Angle	(300A,0125)	R+*	Shall be zero.
>>Table Top Eccentric Rotation Direction	(300A,0126)	R+*	Shall be NONE.
>> Table Top Pitch Angle	(300A,0140)	R+*	Shall be zero.
>> Table Top Pitch Rotation Direction	(300A,0142)	R+*	Shall be NONE.
>> Table Top Roll Angle	(300A,0144)	R+*	Shall be zero.
>> Table Top Roll Rotation Direction	(300A,0146)	R+*	Shall be NONE
>> Table Top Vertical Position	(300A,0128)	O+*	If value is present, shall be constant.
>> Table Top Longitudinal Position	(300A,0129)	O+*	If value is present, shall be constant.
>> Table Top Lateral Position	(300A,012A)	O+*	If value is present, shall be constant.

7.4.4.3 Beam Option Specifications

7.4.4.3.1 Bolus Beam Modifier Base Content

850 7.4.4.3.1.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.3.1.2 Required Attributes

One or more Boli may be optionally included in any of the Treatment Planning - Plan Content Storage and Retrieval transactions (Producer and Consumer Actors).

855 For actors supporting the Bolus Beam Modifier, the attributes specified in the table below have these additional requirements if Number of Boli (300A,00ED) is greater than zero.

Attribute	Tag	Beam Modifier	
		Bolus	
		Presence	Specific Rules
> Number of Boli	(300A,00ED)	R+*	Shall be >=1.
> Referenced Bolus Sequence	(300A,00B0)	R+*	
>> Bolus ID	(300A,00DC)	R+*	Shall be present.

7.4.4.3.2 Block Beam Modifier Base Content

860 7.4.4.3.2.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.3.2.2 Required Attributes

One or more Blocks may be optionally included in some of the Treatment Planning - Plan Content Storage and Retrieval transactions (Producer and Consumer Actors):

865 For actors supporting the Block Beam Modifier, the attributes specified in the table below have these additional requirements if Number of Blocks (300A,00F0) is greater than zero.

Attribute	Tag	Beam Modifier	
		Block	
		Presence	Specific Rules
> Number of Blocks	(300A,00F0)	R+*	Photon Beams: Shall be 0 - 8. Electron Beams: Shall be 0 or 1.
> Block Sequence	(300A,00F4)	R+*	
>> Block Tray ID	(300A,00F5)	R+	See Note 1.
>> Source to Block Tray Distance	(300A,00F6)	R+	
>> Block Divergence	(300A,00FA)	R+*	
>> Block Mounting Position	(300A,00FB)	R+	Shall be present, and shall be handled safely for enumerated values not supported.
>> Material ID	(300A,00E1)	R+	
>> Block Thickness	(300A,0100)	R+	
>> Block Number of Points	(300A,0104)	R+*	
>> Block Data	(300A,0106)	R+*	

Note 1:

870 Photon Beams: There may be multiple blocks with the same Block Tray ID (i.e., placed on the same Tray, e.g., an aperture block and a shield block). If/when Treatment Delivery Verification is taking place, it is essentially the Block Tray that is verified, not the individual blocks on the tray, so one would expect the same Accessory ID to be reused for the same Block Tray ID. This is why Block Name is not made Mandatory (R+*) for the profile. Only a single Block Tray ID shall be supported for a given beam.

Electron Beams: The Block Tray ID defines the electron insert which is checked by the TDD.

875 7.4.4.3.3 Compensator Beam Modifier Base Content

7.4.4.3.3.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.3.3.2 Required Attributes

880 A single Compensator may be optionally included in some of the Treatment Planning - Plan Content Storage and Retrieve transactions (Producer and Consumer Actors):

For actors supporting the Compensator Beam Modifier, the attributes specified in the table below have these additional requirements: if Number of Compensators (300A,00E0) is equal to one.

Attribute	Tag	Beam Modifier	
		Compensator	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	
> Number of Compensators	(300A,00E0)	R+*	Shall be 1.
> Compensator Sequence	(300A,00E3)	R+*	
>> Compensator Type	(300A,00EE)	R+*	Shall be STANDARD.
>> Material ID	(300A,00E1)	R+*	
>> Compensator ID	(300A,00E5)	R+*	
>> Source to Compensator Tray Distance	(300A,00E6)	R+*	
>> Compensator Divergence	(300A,02E0)	R+*	
>> Compensator Mounting Position	(300A,02E1)	R+*	Shall be PATIENT_SIDE or SOURCE_SIDE.
>> Compensator Transmission Data	(300A,00EB)	R+*	
>> Compensator Thickness Data	(300A,00EC)	R+*	

7.4.4.3.4 Hard Wedge Beam Modifier Base Content

885 7.4.4.3.4.1 Referenced Standards

DICOM 2018e Edition PS 3.3

7.4.4.3.4.2 Required Attributes

A single Hard Wedge may be optionally included in some of the Treatment Planning - Plan Content Storage and Retrieve transactions (Producer and Consumer Actors):

890 For actors supporting the Hard Wedge Beam Modifier, the attributes specified in the table below have these additional requirements if Number of Wedges (300A,00D0) is greater than 0.

Attribute	Tag	Beam Modifier	
		Hard Wedge	
		Presence	Specific Rules
Beam Sequence	(300A,00B0)	R+*	

Attribute	Tag	Beam Modifier	
		Hard Wedge	
		Presence	Specific Rules
> Number of Wedges	(300A,00D0)	R+*	Shall be 1 or 2; if 2, one will be Hard and the other will be either Motorized or Virtual
> Wedge Sequence	(300A,00D1)	R+*	
>> Wedge Type	(300A,00D3)	R+*	Shall be STANDARD.
>> Wedge ID	(300A,00D4)	R+	
>> Wedge Angle	(300A,00D5)	R+	
>> Wedge Orientation	(300A,00D8)	R+	
>> Source to Wedge Tray Distance	(300A,00DA)	R+	
...			
> Control Point Sequence	(300A,0111)	R+*	
>> Wedge Position Sequence	(300A,0116)	R+*	If present may not be ignored.
>>> Wedge Position	(300A,0118)	R+*	Shall be IN.

7.4.4.4 Other RT Beam Modules

895 7.4.4.4.1 RT Beams Module for Geometric Planner

7.4.4.4.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.4.4.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Beam Sequence	(300A,00B0)	R+*	An actor must be able to safely handle up to 100 Beam Sequence Items (beams).
>Beam Name	(300A,00C2)	R+	The Beam Name must be unique within the sequence.
>Beam Type	(300A,00C4)	R+*	In the BRTO Profile, for Geometric Plans the value shall be STATIC. Only static beams shall be specified in Geometric Plans. This will allow non-arc-based IMRT (such as Step-and-Shoot or Sliding Window techniques, but not techniques such as fixed aperture arc beams, conformal arc beams, or intensity modulated arc beams. As a result, all beams in Geometric Plans shall consist of exactly two control points.
>Radiation Type	(300A,00C6)	R+*	Any value other than PHOTON is outside the scope of the profile

Attribute	Tag	Type	Attribute Note
>High-Dose Technique Type	(300A,00C7)	O+*	Geometric Plans shall not specify this attribute.
>Treatment Machine Name	(300A,00B2)	O+*	An actor must not rely on the presence of this attribute.
>Source-Axis Distance	(300A,00B4)	R+*	This attribute is critical for providing information regarding beam divergence.
>Beam Limiting Device Sequence	(300A,00B6)		For the BRTO Profile, shall report at least one set of MLC descriptions or the descriptions of two sets of jaws.
>Referenced Patient Setup Number	(300C,006A)	R+*	
>Number of Wedges	(300A,00D0)	R+*	Geometric Plans are constrained to a value of 0 (i.e., a Geometric Plan must not include a Wedge).
>Number of Compensators	(300A,00E0)	R+*	Geometric Plans are constrained to a value of 0 (i.e., a Geometric Plan must not include a Compensator).
>Number of Boli	(300A,00ED)	R+*	Geometric Plans are constrained to a value of 0 (i.e., a Geometric Plan must not include any Boli).
>Number of Blocks	(300A,00F0)	R+*	All actors shall be able to handle 8 block items, of which no more than one may be an aperture
>Block Sequence	(300A,00F4)		
>>Block Divergence	(300A,00FA)	R+*	Must be present and non-null if Block Sequence is present (i.e., when Number of Blocks is 1 or more), with a value of PRESENT
>>Block Number of Points	(300A,0104)	R+*	The value is constrained to be 3 or more.
>>Block Data	(300A,0106)	R+*	Shall be present and non-null. Limitations on the total number of points are limited only by DICOM limitations on representation with 'explicit VR' in total byte lengths. Systems that limit support of legal sequences shall safely handle receipt of such sequences that exceed their limitations, and document this behavior in their IHE-RO Profile adherence statement.
>Applicator Sequence	(300A,0107)		Not expected in Geometric Plans. However, if present, shall be handled in a safe manner by the receiving system (and document this behavior in their IHE-RO Profile adherence statement). Applications exporting this sequence are outside the scope of the BRTO Profile.
>Final Cumulative Meterset Weight	(300A,010E)	O+*	Shall not be present in a Geometric Plan.
>Number of Control Points	(300A,0110)	R+*	Shall have a value of 2 for Geometric Plans.
>Control Point Sequence	(300A,0111)	R+*	In the BRTO Profile, for Geometric Plans the second control point (sequence item) shall contain only: <ul style="list-style-type: none"> Control Point Index (300A,0112) with a value of 1 Cumulative Meterset Weight (300A,0134) set to NULL.
>>Cumulative Meterset Weight	(300A,0134)	O+*	Shall be NULL for Geometric Plans (in both the first and second control point).

Attribute	Tag	Type	Attribute Note
>>Referenced Dose Reference Sequence	(300C,0050)	O+*	Shall not be present for Geometric Plans. Must not be relied upon by actors operating on the object as a Geometric Plan.
>>Nominal Beam Energy	(300A,0114)	O+*	Actors must not rely on the presence of this attribute to operate correctly. However, if this attribute is present, actors may not ignore the value.
>>Dose Rate Set	(300A,0115)	O+*	Actors must not rely on the presence of this attribute to operate correctly. However, if this attribute is present, actors may not ignore the value.
>>Wedge Position Sequence	(300A,0116)	O+*	Must not be present in a Geometric Plan
>>Beam Limiting Device Position Sequence	(300A,011A)	R+*	Must be present and correspond to those devices defined in the Beam Limiting Device Sequence. It shall be present for a Geometric Plan for Control Point Index 0 only.
>>Gantry Rotation Direction	(300A,011F)	R+*	For a Geometric Plan for Control Point Index 0 only, must have a value of NONE.

900 **7.4.5 Plan-Related Modules in Delivery**

This section is present only to convey the envisioned section numbering.

7.4.6 Image-Related Modules in Planning

7.4.6.1 RT Image Module

This section is present only to convey the envisioned section numbering.

905 **7.4.6.2 Image Plane Module**

7.4.6.2.1 Image Plane Base Content

7.4.6.2.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.6.2.1 Module Definition

Attribute	Tag	Type	Attribute Note
Image Orientation (Patient)	(0020,0037)	R+*	This element shall be restricted to TRANSVERSE images only. For a transverse image, direction cosines shall be ($\pm 1, 0, 0, 0, \pm 1, 0$) with an angle tolerance of 0.001 radians (~ 0.057 degrees)
Slice Thickness	(0018,0050)	-	Shall not be relied on.
Slice Location	(0020,1041)	-	Shall not be relied on.
Pixel Spacing	(0028,0030)	O+*	For CT, non-isotropic pixels are outside the scope of the profile. For RT Dose, pixel spacing may be non-isotropic.

910 **7.4.6.2.2 Image Plane Decubitus****7.4.6.2.2.1 Referenced Standards**

DICOM 2018d Edition PS 3.3

7.4.6.2.2.2 Module Definition

Attribute	Tag	Type	Attribute Note
Image Orientation (Patient)	(0020,0037)	R+*	This element shall be restricted to TRANSVERSE images only. For a transverse image, direction cosines shall be ($\pm 1, 0, 0, 0, \pm 1, 0$) or ($0, \pm 1, 0, \pm 1, 0, 0$).with an angle tolerance of 0.001 radians (~ 0.057 degrees)
Slice Thickness	(0018,0050)	-	Shall not be relied on.
Slice Location	(0020,1041)	-	Shall not be relied on.
Pixel Spacing	(0028,0030)	O+*	For CT, non-isotropic pixels are outside the scope of the profile. For RT Dose, pixel spacing may be non-isotropic.

7.4.7 Image-Related Modules in Delivery915 *This section is present only to convey the envisioned section numbering.***7.4.8 Segment-Related Modules****7.4.8.1 RT ROI Observation Module****7.4.8.1.1 RT ROI Observation Module Base Content****7.4.8.1.1.1 Referenced Standards**

920 DICOM 2018d Edition PS 3.3

7.4.8.1.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
RT ROI Observations Sequence	(3006,0080)	R+*	This sequence contains information about an ROI. It references the ROI in Referenced ROI Number which contains a number which must match one of the ROI numbers in one of the elements of the Structure Set ROI Sequence.
>Referenced ROI Number	(3006,0084)	R+*	Specifies the ROI to which this observation applies. For every item in Structure Set ROI sequence, at least one observation is required, with values in ROI Interpreted Type.
>RT ROI Interpreted Type	(3006,00A4)	R+*	If referenced ROI has associated contours of type CLOSED_PLANAR, the content consumer must accept at minimum the following values: EXTERNAL PTV CTV GTV TREATED_VOLUME IRRAD_VOLUME BOLUS AVOIDANCE ORGAN MARKER CONTRAST_AGENT CAVITY If referenced ROI has associated contours of type POINT, the content consumer must accept at minimum the following values: MARKER REGISTRATION ISOCENTER
> Segmented Property Category Code Sequence	(0062,0003)	-	See Note 1
> RT ROI Identification Code Sequence	(3006,0086)	-	See Note 2
>>Segmented Property Type Modifier Code Sequence	(0062,0011)	O+	Not required; Shall contain only one code if present.
>ROI Physical Properties Sequence	(3006,00B0)	O+*	Not required, but shall not be ignored if supplied.
>>ROI Physical Property	(3006,00B2)	R+*	Only relative electron density shall be supported: REL_ELEC_DENSITY

Note 1: This attribute allows preserving information by copying the content of Segmented Property Category Code Sequence (0062,0003) in case a Segmentation object is re-encoded as an RT Structure Set or vice-versa.

925

Note 2: In case of re-encoding a Segmentation object as an RT Structure Set or vice-versa it is suggested that the Segmented Property Type Code Sequence (0062,000F) is mapped to RT ROI Identification Code Sequence (3006,0086).

7.4.8.2 RT ROI Contour Module

7.4.8.2.1 RT ROI Contour Module Base Content

7.4.8.2.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

930

7.4.8.2.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
ROI Contour Sequence	(3006,0039)	R*	
>ROI Display Color	(3006,002A)	-	Not required - no compliant implementation shall rely on this element being present for proper operation. However applications are allowed to be aware of this element and use it to map display colors.
> Recommended Display Grayscale Value	(0062,000C)	-	Not required - no compliant implementation shall rely on this element being present for proper operation. However applications are allowed to be aware of this element and use it to map display colors.
> Recommended Display CIELab Value	(0062,000D)	-	Not required - no compliant implementation shall rely on this element being present for proper operation. However applications are allowed to be aware of this element and use it to map display colors.
>Contour Sequence	(3006,0040)	R+*	Shall be present. Shall contain an item for each contour in the ROI. Compliant implementations shall be able to handle as many as 1000 contours on a single slice. That is, the number of contours in items in all Contour Sequences with the same z-coordinate (and referenced CT image) should be less than or equal to 1000.
>>Contour Image Sequence	(3006,0016)	R+*	Shall be present with a single item. This item is the image upon which this contour should be placed. If the contour type is CLOSED_PLANAR, then the z-coordinates of the contour shall match the z-coordinate of Image Position (Patient) in the image.
>>>Referenced SOP Class UID	(0008,1150)	R+*	Shall be present with a value of '1.2.840.10008.5.1.4.1.1.2'
>>>Referenced SOP Instance UID	(0008,1155)	R*	SOP Instance UID of the image being referenced.
>>>Referenced Frame Number	(0008,1160)	O+*	Shall not be present
>>Contour Geometric Type	(3006,0042)	R+*	Shall be present, with a value of POINT or CLOSED_PLANAR. Conforming implementations must properly interpret this value.

Attribute	Tag	Type	Attribute Note
>>Contour Slab Thickness	(3006,0044)	-	Not required - no compliant implementation shall rely on this element being present for proper operation.
>>Contour Offset Vector	(3006,0045)	O+*	The profile requires that this attribute be zero if present.
>>Number of Contour Points	(3006,0046)	R+*	Required, and must match the actual number of points in Contour Data. Shall not exceed the number for which the Contour Data cannot be encoded when using explicit transfer syntax.
>>Contour Data	(3006,0050)	R+*	Shall be present. If contour type is CLOSED_PLANAR, then all points must have the same z-coordinate. This z-coordinate shall match the z-coordinate in the related CT image within 0.01 mm (contained in the Contour Image sequence in the same item of the ROI Contour Sequence as this data). An implication of this is that the CLOSED_PLANAR contours are transverse.

7.4.8.2.2 RT ROI Contour Module Off-slice

7.4.8.2.2.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.8.2.2.2 Module Definition

Attribute	Tag	Type	Attribute Note
ROI Contour Sequence	(3006,0039)	R*	
>ROI Display Color	(3006,002A)	-	Not required - no compliant implementation shall rely on this element being present for proper operation. However applications are allowed to be aware of this element and use it to map display colors.
> Recommended Display Grayscale Value	(0062,000C)	-	Not required - no compliant implementation shall rely on this element being present for proper operation. However applications are allowed to be aware of this element and use it to map display colors.
> Recommended Display CIELab Value	(0062,000D)	-	Not required - no compliant implementation shall rely on this element being present for proper operation. However applications are allowed to be aware of this element and use it to map display colors.
>Contour Sequence	(3006,0040)	R+*	Shall be present. Shall contain an item for each contour in the ROI. Compliant implementations shall be able to handle as many as 1000 contours on a single slice. That is, the number of contours in items in all Contour Sequences with the same z-coordinate (and referenced CT image) should be less than or equal to 1000.
>> Contour Number	(3006,0048)	R+*	Shall be present if Contour Geometry Type (3006,0042) is CLOSED_PLANAR.
>> Attached Contours	(3006,0049)	RC+*	Shall be present if Contour Geometry Type (3006,0042) is CLOSED_PLANAR and there are other contours referenced.

Attribute	Tag	Type	Attribute Note
			Multiplicity equals the number of contours referenced from this contour (s. 3.3.4.1.2).
>>Contour Image Sequence	(3006,0016)	RC+*	Shall be present for contours located on image planes. This item is the image upon which this contour should be placed. If the contour type is CLOSED_PLANAR, there shall be contours whose z-coordinates match the z-coordinates of Image Position (Patient) in the image for structures that intersect this image plane.
>>>Referenced SOP Class UID	(0008,1150)	R+*	Shall be present with a value of '1.2.840.10008.5.1.4.1.1.2'
>>>Referenced SOP Instance UID	(0008,1155)	R*	SOP Instance UID of the image being referenced.
>>>Referenced Frame Number	(0008,1160)	O+*	Shall not be present
>>Contour Geometric Type	(3006,0042)	R+*	Shall be present, with a value of POINT or CLOSED_PLANAR. Conforming implementations must properly interpret this value.
>>Contour Slab Thickness	(3006,0044)	-	Not required - no compliant implementation shall rely on this element being present for proper operation.
>>Contour Offset Vector	(3006,0045)	O+*	The profile requires that this attribute be zero if present.
>>Number of Contour Points	(3006,0046)	R+*	Required, and must match the actual number of points in Contour Data. Shall not exceed the number for which the Contour Data cannot be encoded when using explicit transfer syntax.
>>Contour Data	(3006,0050)	R+*	Shall be present. If contour type is CLOSED_PLANAR, then all points must have the same z-coordinate. For every image plane which is referenced in the Structure Set Module () and intersect the ROI, there shall be contours defined the image plane. The z-coordinate of those contours shall match the z-coordinate of the referenced image plane within 0.01 mm (contained in the Contour Image sequence in the same item of the ROI Contour Sequence as this data). An implication of this is that the CLOSED_PLANAR contours are transverse.

935 7.4.8.3 RT Structure Set Module

7.4.8.3.1 Structure Set Module Base Content

7.4.8.3.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.8.3.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Structure Set Label	(3006,0002)	R+	
Structure Set Date	(3006,0008)	R+	
Structure Set Time	(3006,0009)	R+	
Referenced Frame of Reference Sequence	(3006,0010)	R+*	This element is required for all 3D RT Structure Sets which are image based. It is to contain a set of references to the entire set of images which comprise the volume from which the Structure Set was constructed, and which is to be used for planning. There should only be one item in this sequence, as a BRTO Profile-based structure is based on a single set of images, which are all in the same frame of reference.
>Frame of Reference UID	(0020,0052)	R+*	This frame of reference UID shall be the same as the frame of reference of the CT series from which the Structure Set was constructed. It will also be the same as the frame of reference of any related RTPLAN's or RTDOSE's.
>RT Referenced Study Sequence	(3006,0012)	R+*	Shall be present and contain the series sequence. Only one item allowed in this sequence.
>>Referenced SOP Instance UID	(0008,1155)	R+*	This Study Instance UID shall be the same as the Study Instance UID of the related CT instances.
>>RT Referenced Series Sequence	(3006,0014)	R+*	Shall be present to contain the Contour Image Sequence. Only one item allowed in this sequence.
>>>Series Instance UID	(0020,000E)	R+*	Shall be present and contain the series to which the set of CT images upon which the structure set is based belong.
>>>Contour Image Sequence	(3006,0016)	R+*	Shall be present. Contains an item for each CT image in the volume upon which the Structure Set is based.
>>>>Referenced SOP Class UID	(0008,1155)	R+*	Shall be present with a value of '1.2.840.10008.5.1.4.1.1.2' This profile is for volumes based on CT Images only
>>>>Referenced Frame Number	(0008,1160)	O+*	Shall not be present
Structure Set ROI Sequence	(3006,0020)	R+*	This sequence shall be present. It defines the ROI's in this Structure Set.
>ROI Number	(3006,0022)	R*	This defines an index to be used for referencing a particular ROI item from other sequences. It is required to be unique within the Structure Set in which it is created. No limitation on values other than uniqueness within sequence.
>Referenced Frame of Reference UID	(3006,0024)	R*	This frame of reference UID shall be the same as the frame of reference UID of the CT series from which the Structure Set was constructed. It will also be the same as the frame of reference of any related RTPLAN or RTDOSE instances.
>ROI Name	(3006,0026)	R+	This is the primary identifier for an ROI (from user perspective). Shall be present and should match UI display. Shall be unique within the Structure Set ROI sequence.
>ROI Description	(3006,0028)	-	Not required - no compliant implementation shall rely on this element being present for proper operation.
>ROI Volume	(3006,002C)	-	Not required - no compliant implementation shall rely on this element being present for proper operation.

Attribute	Tag	Type	Attribute Note
>ROI Generation Algorithm	(3006,0036)	R+	<p>Shall be present, with a value of AUTOMATIC, SEMIAUTOMATIC, or MANUAL.</p> <p>This information may be presented to a user, but no semantics for handling a Structure Set is required for this profile.</p> <p>Implementations which create Structure Set instances must provide an appropriate value.</p>

940 7.4.8.3.2 Structure Set in Multi Modality

7.4.8.3.2.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.8.3.2.2 Module Definition

945 The following table lists redefinitions of attributes within the Multimodality Image Registration for Radiation Oncology Integration Profile, which extend the definition of Structure Set Base Content (see 7.4.8.3.1). Attributes displayed in a light grey value are not modified but only added to provide the context in which a certain attribute enhancement is defined.

Attribute	Tag	Type	Attribute Note
Referenced Frame of Reference Sequence	(3006,0010)	R+*	This element is required for all 3D RT Structure Sets which are image based. It is to contain a set of references to the entire set of images which comprise the volume from which the Structure Set was constructed, and which is to be used for planning. There should only be one item in this sequence, as a BRTO Profile-based structure is based on a single set of images, which are all in the same frame of reference.
>RT Referenced Study Sequence	(3006,0012)	R+*	Shall be present and contain the series sequence. Only one item allowed in this sequence.
>>RT Referenced Series Sequence	(3006,0014)	R+*	Shall be present to contain the Contour Image Sequence. Only one item allowed in this sequence.
>>>Contour Image Sequence	(3006,0016)	R+*	Shall be present. Contains an item for each CT image in the volume upon which the Structure Set is based.
>>>>Referenced SOP Class UID	(0008,1155)	R+*	Must be present with a value of enhancement '1.2.840.10008.5.1.4.1.1.2', '1.2.840.10008.5.1.4.1.1.4' or '1.2.840.10008.5.1.4.1.1.128'
Structure Set ROI Sequence	(3006,0020)	R+*	This sequence shall be present. It defines the ROI's in this Structure Set.
>ROI Generation Algorithm	(3006,0036)	R+	<p>Must be present, with a value of AUTOMATIC, SEMIAUTOMATIC, MANUAL, or RESAMPLED.</p> <p>This information may be presented to a user, but no semantics for handling an RTSTRUCT is required for this profile.</p>

Attribute	Tag	Type	Attribute Note
			RESAMPLED indicates that the ROI Contours have been resampled onto a different set of images from those on which the contours were originally created. Implementations which create RTSTRUCT instances must provide an appropriate value.

950 7.4.9 Segment Modules in Delivery

This section is present only to convey the envisioned section numbering.

7.4.10 Registration Modules in Planning

7.4.10.1 Registration Module

7.4.10.1.1 Registration Module Base Content

955 7.4.10.1.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.10.1.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Content Label	(0070,0080)	-	
Content Description	(0070,0081)	R+	Description used to distinguish registration instances. Shall not be empty.
Registration Sequence	(0070,0308)	R	The sequence shall contain 2 registration items. One Frame of Reference will be to the Registered Frame of Reference, the other will define the spatial registration from the specified Frame of Reference to the Registered Frame of Reference. Note: The order of these items is not significant
>Frame of Reference UID	(0020,0052)	R*	Identifies a Frame of Reference that may or may not be an image set (e.g., atlas or physical space). See Section C.7.4.1.1.1 for further explanation. Shall be present.
>Referenced Image Sequence	(0008,1140)	R*	Identifies the set of images registered in this sequence item. One or more items shall be present. Represents the set of images available to the Registrator at the time of spatial registration (see RO TF-2: 3.17.4.1.2).
>Matrix Registration Sequence	(0070,0309)	R	A sequence that specifies one spatial registration. Exactly one item shall be present
>>Matrix Sequence	(0070,030A)	R	One item shall be present. The item specifies a transformation. See Section C.20.2.1.1 in the DICOM Standard

Attribute	Tag	Type	Attribute Note
>>>Frame of Reference Transformation Matrix	(3006,00C6)	R	A 4x4 homogeneous transformation matrix that registers the referenced images to the Registered Frame of Reference. Matrix elements shall be listed in row-major order. See Section C.20.2.1.1 in the DICOM Standard
>>>Frame of Reference Transformation Matrix Type	(0070,030C)	R	The only type of Frame of Reference Transformation Matrix (3006,00C6) supported in this profile is RIGID. See Section C.20.2.1.2 in the DICOM Standard.

7.4.11 Treatment Records

This section is present only to convey the envisioned section numbering.

960 7.4.12 Prescription-Related Modules in Planning

This section is present only to convey the envisioned section numbering.

7.4.13 Dose-Related Modules

7.4.13.1 Image Plane Module

7.4.13.1.1 Image Plane Base Content

965 7.4.13.1.1.1 Referenced Standards

DICOM 2018d Edition PS 3.3

7.4.13.1.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Image Orientation (Patient)	(0020,0037)	R+*	This element shall be restricted to TRANSVERSE images only. For a transverse image, direction cosines shall be ($\pm 1, 0, 0, 0, \pm 1, 0$) with an angle tolerance of 0.001 radians (~ 0.057 degrees). The Image Orientation (Patient) shall correspond to the RT Patient Setup of the associated RT Plan (Section 7.3.2.2.1)
Slice Thickness	(0018,0050)	-	Shall not be relied on.
Slice Location	(0020,1041)	-	Shall not be relied on.
Pixel Spacing	(0028,0030)	O+*	For CT, non-isotropic pixels are outside the scope of the profile. For RT Dose, pixel spacing may be non-isotropic.

7.4.13.2 Multi-Frame Module970 **7.4.13.2.1 Multi-Frame Module Base Content****7.4.13.2.1.1 Referenced Standards**

DICOM 2018d Edition PS 3.3

7.4.13.2.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Frame Increment Pointer	(0028,0009)	R+*	Shall have the same value as the Grid Frame Offset Vector (3004,000C).

7.4.13.3 RT Dose Module975 **7.4.13.3.1 RT Dose Module Base Content****7.4.13.3.1.1 Referenced Standards**

DICOM 2018d Edition PS 3.3

7.4.13.3.1.2 Module Definition

Attribute	Tag	Type	Attribute Note
Content Date	(0008,0023)	R+	Required
Content Time	(0008,0033)	R+	Required
Samples per Pixel	(0028,0002)	R+*	Shall be present and equal to 1.
Photometric Interpretation	(0028,0004)	R+*	Shall be present and equal to MONOCHROME2 .
Bits Allocated	(0028,0100)	R+*	Shall be present and equal to 16 or 32.
Bits Stored	(0028,0101)	R+*	Shall be equal to Bits Allocated.
High Bit	(0028,0102)	R+*	Shall be one less than Bits Stored.
Pixel Representation	(0028,0103)	R+*	Shall have the value 0 = unsigned integer. Negative dose values shall not be present.
Dose Units	(3004,0002)	R+*	Shall be equal to the enumerated value GY .
Dose Type	(3004,0004)	R+*	Shall be equal to the defined term PHYSICAL or EFFECTIVE .
Dose Comment	(3004,0006)	RC+	Shall be present and not empty if Referenced RT Plan Sequence (300C,0002) is missing and RT Plan Description is present, in which case it should have the same value as RT Plan Description.
Normalization Point	(3004,0008)	-	Shall not be relied on.
Dose Summation Type	(3004,000A)	R+*	Shall have the value PLAN .
Referenced RT Plan Sequence	(300C,0002)	R+*	Shall be present if Dose Summation Type (3004,000A) has the value PLAN .

Attribute	Tag	Type	Attribute Note
Grid Frame Offset Vector	(3004,000C)	R+*	First z coordinate shall be equal to zero. The remaining z coordinates shall be relative to the starting z position in Image Position (Patient) (0020,0032). The difference between neighboring values shall be constant with a tolerance of 0.01mm.
Tissue Heterogeneity Correction	(3004,0014)	R+	Shall be present.

7.4.13.4 RT DVH Module

980 7.4.13.4.1 RT DVH Module Base Content

7.4.13.4.1.1 Referenced Standard

DICOM 2018d Edition PS 3.3

7.4.13.4.1.2 Module Definition

Attribute Name	Tag	Type	Attribute Description
DVH Normalization Point	(3004,0040)	R+*	Shall not be present
DVH Normalization Dose Value	(3004,0042)	R+*	Shall not be present
DVH Sequence	(3004,0050)	-	Sequence of DVHs. One or more Items shall be included in this Sequence.
>DVH Type	(3004,0001)	R+*	Shall be DIFFERENTIAL or CUMULATIVE
>Dose Units	(3004,0002)	R+	Shall be GY.
>Dose Type	(3004,0004)	R+	Shall be either PHYSICAL or EFFECTIVE
>DVH Volume Units	(3004,0054)	R+	Shall be CM3

8 IHE RO Content Modules (type other than CDA or DICOM)

The structure for other types of “Content” sections has not yet been defined. Authors utilizing this section are encouraged to provide input to development of the template structure for this Technical Framework Volume 3 section at http://ihe.net/Templates_Public_Comments.

990

Appendices

None

995 **Glossary**

IHE Glossary terms are maintained via the [Standards Knowledge Management Tool](#) (SKMT). SKMT contains IHE Glossary Terms from Final Text and Trial Implementation IHE profiles. Terms from Public Comment profiles are added once the profile is published for Trial Implementation.

1000 The current published version of the IHE Glossary ([Appendix D](#) to the *IHE Technical Frameworks General Introduction*) contains IHE Glossary terms as of the publication date of the document. Always refer to the [SKMT](#) for the most up-to-date list of IHE Glossary terms.

Further information about SKMT can be found on the IHE Wiki [here](#).

1005